

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF062419\
 Data File : BF115160.D
 Acq On : 24 Jun 2019 21:15
 Operator : HP/JU
 Sample : K3500-05
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 B-5(9-10)

Quant Time: Jun 25 08:21:10 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF062119.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 24 04:21:29 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.61	152	252390	20.00	ng	0.00
21) Naphthalene-d8	7.89	136	959343	20.00	ng	-0.01
39) Acenaphthene-d10	9.64	164	475905	20.00	ng	-0.01
64) Phenanthrene-d10	11.11	188	955896	20.00	ng	-0.01
76) Chrysene-d12	13.72	240	721961	20.00	ng	-0.02
87) Perylene-d12	15.09	264	618886	20.00	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.21	112	1392080	85.11	ng	0.00
7) Phenol-d6	6.25	99	1761422	88.73	ng	0.00
23) Nitrobenzene-d5	7.17	82	1043725	66.93	ng	-0.01
42) 2,4,6-Tribromophenol	10.42	330	445548	88.78	ng	-0.01
45) 2-Fluorobiphenyl	8.97	172	1837766	65.59	ng	-0.02
79) Terphenyl-d14	12.69	244	2073480	61.06	ng	-0.01
Target Compounds						
50) Dimethylphthalate	9.36	163	220366	6.294	ng	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF062419\
Data File : BF115160.D
Acq On : 24 Jun 2019 21:15
Operator : HP/JU
Sample : K3500-05
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
B-5(9-10)

Quant Time: Jun 25 08:21:10 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF062119.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 24 04:21:29 2019
Response via : Initial Calibration

