

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF062419\
 Data File : BF115166.D
 Acq On : 24 Jun 2019 23:54
 Operator : HP/JU
 Sample : K3500-08
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 B-8(9-10)

Quant Time: Jun 25 08:34:18 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF062119.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 24 04:21:29 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.61	152	278141	20.00	ng	0.00
21) Naphthalene-d8	7.89	136	1062392	20.00	ng	-0.01
39) Acenaphthene-d10	9.64	164	532671	20.00	ng	-0.01
64) Phenanthrene-d10	11.11	188	1047526	20.00	ng	-0.01
76) Chrysene-d12	13.73	240	712787	20.00	ng	-0.02
87) Perylene-d12	15.09	264	648322	20.00	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.21	112	1323148	73.41	ng	0.00
7) Phenol-d6	6.25	99	1629497	74.49	ng	0.00
23) Nitrobenzene-d5	7.17	82	971501	56.25	ng	-0.01
42) 2,4,6-Tribromophenol	10.42	330	466027	82.96	ng	-0.01
45) 2-Fluorobiphenyl	8.97	172	1806177	57.60	ng	-0.02
79) Terphenyl-d14	12.69	244	1997280	59.58	ng	-0.01
Target Compounds						
50) Dimethylphthalate	9.36	163	182272	4.651	ng	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF062419\
Data File : BF115166.D
Acq On : 24 Jun 2019 23:54
Operator : HP/JU
Sample : K3500-08
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
B-8(9-10)

Quant Time: Jun 25 08:34:18 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF062119.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 24 04:21:29 2019
Response via : Initial Calibration

