

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF062419\
 Data File : BF115171.D
 Acq On : 25 Jun 2019 2:07
 Operator : HP/JU
 Sample : K3500-04
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 B-4(9-10)

Quant Time: Jun 25 08:26:43 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF062119.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 24 04:21:29 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.61	152	279190	20.00	ng	0.00
21) Naphthalene-d8	7.89	136	1064444	20.00	ng	-0.01
39) Acenaphthene-d10	9.64	164	528078	20.00	ng	-0.01
64) Phenanthrene-d10	11.11	188	1058366	20.00	ng	-0.01
76) Chrysene-d12	13.72	240	699709	20.00	ng	-0.02
87) Perylene-d12	15.09	264	631422	20.00	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.21	112	1426958	78.87	ng	0.00
7) Phenol-d6	6.25	99	1795214	81.75	ng	0.00
23) Nitrobenzene-d5	7.17	82	1047293	60.52	ng	-0.01
42) 2,4,6-Tribromophenol	10.42	330	479769	86.15	ng	-0.01
45) 2-Fluorobiphenyl	8.97	172	1756267	56.49	ng	-0.02
79) Terphenyl-d14	12.69	244	1886542	57.32	ng	-0.01
Target Compounds						
50) Dimethylphthalate	9.36	163	370997	9.549	ng	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF062419\
Data File : BF115171.D
Acq On : 25 Jun 2019 2:07
Operator : HP/JU
Sample : K3500-04
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
B-4(9-10)

Quant Time: Jun 25 08:26:43 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF062119.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 24 04:21:29 2019
Response via : Initial Calibration

