

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF062419\
 Data File : BF115153.D
 Acq On : 24 Jun 2019 18:03
 Operator : HP/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 25 08:07:58 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF062119.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 24 04:21:29 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	129	0.00
2	1,4-Dioxane	0.612	0.618	-1.0	127	0.00
3	Pyridine	1.724	1.682	2.4	123	-0.01
4	n-Nitrosodimethylamine	0.676	0.650	3.8	122	0.00
5 S	2-Fluorophenol	1.296	1.292	0.3	124	0.00
6	Aniline	2.162	2.230	-3.1	127	0.00
7 S	Phenol-d6	1.573	1.587	-0.9	125	0.00
8	2-Chlorophenol	1.457	1.456	0.1	124	0.00
9	Benzaldehyde	1.026	0.866	15.6	102	0.00
10 C	Phenol	1.843	1.785	3.1	121	0.00
11	bis(2-Chloroethyl)ether	1.358	1.364	-0.4	126	0.00
12	1,3-Dichlorobenzene	1.617	1.669	-3.2	129	0.00
13 C	1,4-Dichlorobenzene	1.622	1.685	-3.9	128	0.00
14	1,2-Dichlorobenzene	1.502	1.554	-3.5	127	0.00
15	Benzyl Alcohol	1.145	1.155	-0.9	123	0.00
16	2,2'-oxybis(1-Chloropropane	1.970	1.996	-1.3	127	0.00
17	2-Methylphenol	1.203	1.194	0.7	125	0.00
18	Hexachloroethane	0.585	0.600	-2.6	128	0.00
19 P	n-Nitroso-di-n-propylamine	1.007	0.969	3.8	122	0.00
20	3+4-Methylphenols	1.389	1.378	0.8	124	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	129	0.00
22	Acetophenone	0.458	0.465	-1.5	127	0.00
23 S	Nitrobenzene-d5	0.325	0.346	-6.5	133	0.00
24	Nitrobenzene	0.358	0.367	-2.5	127	0.00
25	Isophorone	0.666	0.649	2.6	124	0.00
26 C	2-Nitrophenol	0.177	0.193	-9.0	135	0.00
27	2,4-Dimethylphenol	0.290	0.287	1.0	124	0.00
28	bis(2-Chloroethoxy)methane	0.421	0.426	-1.2	126	0.00
29 C	2,4-Dichlorophenol	0.299	0.302	-1.0	125	0.00
30	1,2,4-Trichlorobenzene	0.330	0.347	-5.2	129	0.00
31	Naphthalene	0.982	0.989	-0.7	124	0.00
32	Benzoic acid	0.207	0.218	-5.3	154#	0.00
33	4-Chloroaniline	0.449	0.465	-3.6	129	0.00
34 C	Hexachlorobutadiene	0.181	0.193	-6.6	132	0.00
35	Caprolactam	0.100	0.104	-4.0	132	0.00
36 C	4-Chloro-3-methylphenol	0.303	0.304	-0.3	125	0.00
37	2-Methylnaphthalene	0.662	0.675	-2.0	126	0.00
38	1-Methylnaphthalene	0.632	0.636	-0.6	123	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	132	0.00
40	1,2,4,5-Tetrachlorobenzene	0.565	0.593	-5.0	130	0.00
41 P	Hexachlorocyclopentadiene	0.335	0.344	-2.7	129	-0.01
42 S	2,4,6-Tribromophenol	0.211	0.227	-7.6	135	0.00
43 C	2,4,6-Trichlorophenol	0.436	0.454	-4.1	130	0.00
44	2,4,5-Trichlorophenol	0.449	0.469	-4.5	134	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF062419\
 Data File : BF115153.D
 Acq On : 24 Jun 2019 18:03
 Operator : HP/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 25 08:07:58 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF062119.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 24 04:21:29 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45 S	2-Fluorobiphenyl	1.177	1.213	-3.1	127	0.00
46	1,1'-Biphenyl	1.600	1.598	0.1	128	0.00
47	2-Chloronaphthalene	1.298	1.334	-2.8	128	0.00
48	2-Nitroaniline	0.378	0.410	-8.5	136	0.00
49	Acenaphthylene	2.022	1.992	1.5	123	0.00
50	Dimethylphthalate	1.471	1.506	-2.4	128	0.00
51	2,6-Dinitrotoluene	0.340	0.352	-3.5	131	0.00
52 C	Acenaphthene	1.266	1.286	-1.6	127	0.00
53	3-Nitroaniline	0.415	0.426	-2.7	130	0.00
54 P	2,4-Dinitrophenol	0.150	0.181	-20.7	157#	0.00
55	Dibenzofuran	1.816	1.751	3.6	123	0.00
56 P	4-Nitrophenol	0.332	0.339	-2.1	126	0.00
57	2,4-Dinitrotoluene	0.439	0.472	-7.5	134	0.00
58	Fluorene	1.352	1.241	8.2	125	0.00
59	2,3,4,6-Tetrachlorophenol	0.377	0.387	-2.7	129	0.00
60	Diethylphthalate	1.479	1.515	-2.4	130	0.00
61	4-Chlorophenyl-phenylether	0.645	0.615	4.7	129	-0.01
62	4-Nitroaniline	0.416	0.444	-6.7	135	-0.01
63	Azobenzene	1.382	1.370	0.9	125	-0.01
64 I	Phenanthrene-d10	1.000	1.000	0.0	131	0.00
65	4,6-Dinitro-2-methylphenol	0.100	0.115	-15.0	146	0.00
66 c	n-Nitrosodiphenylamine	0.623	0.629	-1.0	126	-0.01
67	4-Bromophenyl-phenylether	0.217	0.227	-4.6	132	0.00
68	Hexachlorobenzene	0.235	0.250	-6.4	135	0.00
69	Atrazine	0.200	0.195	2.5	122	-0.01
70 C	Pentachlorophenol	0.150	0.160	-6.7	132	-0.01
71	Phenanthrene	1.077	1.036	3.8	125	0.00
72	Anthracene	1.087	1.062	2.3	125	-0.01
73	Carbazole	0.971	0.972	-0.1	125	0.00
74	Di-n-butylphthalate	1.122	1.160	-3.4	129	-0.01
75 C	Fluoranthene	1.074	1.085	-1.0	126	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	126	-0.01
77	Benzidine	0.888	0.795	10.5	109	-0.01
78	Pyrene	1.537	1.555	-1.2	125	-0.01
79 S	Terphenyl-d14	0.941	0.906	3.7	118	-0.01
80	Butylbenzylphthalate	0.678	0.740	-9.1	135	-0.01
81	Benzo(a)anthracene	1.435	1.476	-2.9	126	-0.01
82	3,3'-Dichlorobenzidine	0.518	0.586	-13.1	136	-0.01
83	Chrysene	1.315	1.370	-4.2	128	0.00
84	Bis(2-ethylhexyl)phthalate	0.889	0.977	-9.9	136	-0.01
85 c	Di-n-octyl phthalate	1.493	1.665	-11.5	137	-0.02
86	Indeno(1,2,3-cd)pyrene	1.556	1.702	-9.4	132	-0.02
87 I	Perylene-d12	1.000	1.000	0.0	134	-0.01

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF062419\
Data File : BF115153.D
Acq On : 24 Jun 2019 18:03
Operator : HP/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Jun 25 08:07:58 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF062119.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 24 04:21:29 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.248	1.276	-2.2	129	-0.01
89	Benzo(k)fluoranthene	1.119	1.067	4.6	133	0.00
90 C	Benzo(a)pyrene	1.125	1.138	-1.2	130	-0.01
91	Dibenzo(a,h)anthracene	1.050	1.110	-5.7	131	-0.02
92	Benzo(q,h,i)perylene	1.063	1.109	-4.3	128	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0