

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF062421\
 Data File : BF124441.D
 Acq On : 24 Jun 2021 14:04
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 24 15:25:50 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF062321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 23 18:48:27 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	101	0.00
2	1,4-Dioxane	40.000	41.290	-3.2	102	0.00
3	Pyridine	40.000	39.906	0.2	93	0.00
4	n-Nitrosodimethylamine	40.000	42.597	-6.5	106	0.00
5 S	2-Fluorophenol	80.000	81.138	-1.4	99	0.00
6	Aniline	40.000	41.278	-3.2	103	0.00
7 S	Phenol-d6	80.000	80.629	-0.8	98	0.00
8	2-Chlorophenol	40.000	40.075	-0.2	100	0.00
9	Benzaldehyde	40.000	42.859	-7.1	101	0.00
10 C	Phenol	40.000	39.704	0.7	103	0.00
11	bis(2-Chloroethyl)ether	40.000	40.795	-2.0	104	0.00
12	1,3-Dichlorobenzene	40.000	39.705	0.7	100	0.00
13 C	1,4-Dichlorobenzene	40.000	40.255	-0.6	103	0.00
14	1,2-Dichlorobenzene	40.000	38.579	3.6	98	0.00
15	Benzyl Alcohol	40.000	40.413	-1.0	104	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	38.373	4.1	97	0.00
17	2-Methylphenol	40.000	39.918	0.2	102	0.00
18	Hexachloroethane	40.000	39.407	1.5	98	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	41.301	-3.3	107	0.00
20	3+4-Methylphenols	40.000	39.115	2.2	96	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	103	0.00
22	Acetophenone	40.000	40.991	-2.5	101	0.00
23 S	Nitrobenzene-d5	80.000	81.300	-1.6	103	0.00
24	Nitrobenzene	40.000	39.009	2.5	105	0.00
25	Isophorone	40.000	39.660	0.9	104	0.00
26 C	2-Nitrophenol	40.000	38.838	2.9	101	0.00
27	2,4-Dimethylphenol	40.000	39.871	0.3	103	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.448	1.4	101	0.00
29 C	2,4-Dichlorophenol	40.000	39.202	2.0	102	0.00
30	1,2,4-Trichlorobenzene	40.000	39.339	1.7	100	0.00
31	Naphthalene	40.000	39.897	0.3	103	0.00
32	Benzoic acid	40.000	35.872	10.3	95	0.00
33	4-Chloroaniline	40.000	39.659	0.9	100	0.00
34 C	Hexachlorobutadiene	40.000	39.028	2.4	105	0.00
35	Caprolactam	40.000	41.421	-3.6	103	0.00
36 C	4-Chloro-3-methylphenol	40.000	39.217	2.0	104	0.00
37	2-Methylnaphthalene	40.000	38.965	2.6	100	0.00
38	1-Methylnaphthalene	40.000	39.470	1.3	103	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	110	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	37.508	6.2	98	0.00
41 P	Hexachlorocyclopentadiene	40.000	23.051	42.4#	107	0.00
42 S	2,4,6-Tribromophenol	80.000	77.001	3.7	104	0.00
43 C	2,4,6-Trichlorophenol	40.000	36.496	8.8	100	0.00
44	2,4,5-Trichlorophenol	40.000	35.575	11.1	96	0.00
45 S	2-Fluorobiphenyl	80.000	75.612	5.5	101	0.00
46	1,1'-Biphenyl	40.000	39.043	2.4	103	0.00
47	2-Chloronaphthalene	40.000	38.531	3.7	107	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF062421\
 Data File : BF124441.D
 Acq On : 24 Jun 2021 14:04
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 24 15:25:50 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF062321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 23 18:48:27 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	38.691	3.3	101	0.00
49	Acenaphthylene	40.000	37.350	6.6	103	0.00
50	Dimethylphthalate	40.000	37.052	7.4	101	0.00
51	2,6-Dinitrotoluene	40.000	35.633	10.9	98	0.00
52 C	Acenaphthene	40.000	37.483	6.3	103	0.00
53	3-Nitroaniline	40.000	38.424	3.9	105	0.00
54 P	2,4-Dinitrophenol	40.000	37.214	7.0	97	0.00
55	Dibenzofuran	40.000	37.005	7.5	102	0.00
56 P	4-Nitrophenol	40.000	34.248	14.4	94	0.00
57	2,4-Dinitrotoluene	40.000	38.285	4.3	105	0.00
58	Fluorene	40.000	37.904	5.2	105	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	38.382	4.0	103	0.00
60	Diethylphthalate	40.000	36.332	9.2	104	0.00
61	4-Chlorophenyl-phenylether	40.000	36.807	8.0	102	0.00
62	4-Nitroaniline	40.000	39.763	0.6	111	0.00
63	Azobenzene	40.000	37.478	6.3	105	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	104	0.00
65	4,6-Dinitro-2-methylphenol	40.000	40.114	-0.3	103	0.00
66 c	n-Nitrosodiphenylamine	40.000	39.919	0.2	106	0.00
67	4-Bromophenyl-phenylether	40.000	38.055	4.9	103	0.00
68	Hexachlorobenzene	40.000	36.601	8.5	101	0.00
69	Atrazine	40.000	43.149	-7.9	105	0.00
70 C	Pentachlorophenol	40.000	35.333	11.7	93	0.00
71	Phenanthrene	40.000	38.270	4.3	104	0.00
72	Anthracene	40.000	37.663	5.8	100	0.00
73	Carbazole	40.000	38.640	3.4	105	0.00
74	Di-n-butylphthalate	40.000	37.803	5.5	104	0.00
75 C	Fluoranthene	40.000	37.094	7.3	103	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	102	0.00
77	Benzydine	40.000	52.063	-30.2#	134	0.00
78	Pyrene	40.000	40.334	-0.8	110	0.00
79 S	Terphenyl-d14	80.000	82.348	-2.9	108	0.00
80	Butylbenzylphthalate	40.000	39.712	0.7	110	0.00
81	Benzo(a)anthracene	40.000	39.985	0.0	107	0.00
82	3,3'-Dichlorobenzidine	40.000	38.479	3.8	107	0.00
83	Chrysene	40.000	39.201	2.0	106	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	38.984	2.5	106	0.00
85 c	Di-n-octyl phthalate	40.000	39.970	0.1	108	0.00
86 I	Perylene-d12	20.000	20.000	0.0	104	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	40.055	-0.1	104	0.00
88	Benzo(b)fluoranthene	40.000	38.801	3.0	107	0.00
89	Benzo(k)fluoranthene	40.000	35.643	10.9	101	0.00
90 C	Benzo(a)pyrene	40.000	38.970	2.6	110	0.00
91	Dibenzo(a,h)anthracene	40.000	39.719	0.7	104	0.00
92	Benzo(g,h,i)perylene	40.000	39.296	1.8	102	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF062421\
Data File : BF124441.D
Acq On : 24 Jun 2021 14:04
Operator : JU/CG
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Jun 24 15:25:50 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF062321.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 23 18:48:27 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0