

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF062422\
 Data File : BF128981.D
 Acq On : 24 Jun 2022 11:01
 Operator : CG\JU
 Sample : PB145742BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB145742BL

Quant Time: Jun 25 06:12:25 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF060222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 24 00:50:48 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.722	152	128231	20.000	ng	0.00
21) Naphthalene-d8	8.004	136	491715	20.000	ng	0.00
39) Acenaphthene-d10	9.763	164	274353	20.000	ng	0.00
64) Phenanthrene-d10	11.251	188	509579	20.000	ng	# 0.00
76) Chrysene-d12	13.898	240	375623	20.000	ng	# 0.00
86) Perylene-d12	15.339	264	288934	20.000	ng	0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.369	112	954009	118.503	ng	0.01
7) Phenol-d6	6.387	99	1199618	122.073	ng	0.00
23) Nitrobenzene-d5	7.292	82	870770	82.450	ng	-0.01
42) 2,4,6-Tribromophenol	10.563	330	338362	104.381	ng	0.00
45) 2-Fluorobiphenyl	9.086	172	1598804	81.632	ng	0.00
79) Terphenyl-d14	12.845	244	1774125	82.071	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF062422\
Data File : BF128981.D
Acq On : 24 Jun 2022 11:01
Operator : CG\JU
Sample : PB145742BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB145742BL

Quant Time: Jun 25 06:12:25 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF060222.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 24 00:50:48 2022
Response via : Initial Calibration

