

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF062422\
 Data File : BF128988.D
 Acq On : 24 Jun 2022 14:33
 Operator : CG\JU
 Sample : N3478-02
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 RO-364

Quant Time: Jun 25 06:13:34 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF060222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 24 00:50:48 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.722	152	104608	20.000	ng	0.00
21) Naphthalene-d8	8.004	136	410721	20.000	ng	0.00
39) Acenaphthene-d10	9.763	164	232144	20.000	ng	0.00
64) Phenanthrene-d10	11.251	188	426356	20.000	ng	# 0.00
76) Chrysene-d12	13.892	240	223641	20.000	ng	# 0.00
86) Perylene-d12	15.333	264	225365	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.351	112	579712	88.271	ng	0.00
7) Phenol-d6	6.381	99	732063	91.318	ng	0.00
23) Nitrobenzene-d5	7.292	82	520974	59.057	ng	-0.01
42) 2,4,6-Tribromophenol	10.557	330	213176	77.719	ng	0.00
45) 2-Fluorobiphenyl	9.086	172	971297	58.610	ng	0.00
79) Terphenyl-d14	12.839	244	922562	71.681	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF062422\
Data File : BF128988.D
Acq On : 24 Jun 2022 14:33
Operator : CG\JU
Sample : N3478-02
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
RO-364

Quant Time: Jun 25 06:13:34 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF060222.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 24 00:50:48 2022
Response via : Initial Calibration

