

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF062422\
 Data File : BF128991.D
 Acq On : 24 Jun 2022 16:00
 Operator : CG\JU
 Sample : N3431-08
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 WC-14B-2-3

Quant Time: Jun 25 06:14:11 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF060222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 24 00:50:48 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.722	152	105704	20.000	ng	0.00
21) Naphthalene-d8	8.004	136	395631	20.000	ng	0.00
39) Acenaphthene-d10	9.763	164	205097	20.000	ng	0.00
64) Phenanthrene-d10	11.251	188	309334	20.000	ng	0.00
76) Chrysene-d12	13.898	240	216586	20.000	ng	# 0.00
86) Perylene-d12	15.333	264	141724	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.369	112	727362	109.605	ng	0.01
7) Phenol-d6	6.381	99	863678	106.618	ng	0.00
23) Nitrobenzene-d5	7.292	82	718122	84.510	ng	-0.01
42) 2,4,6-Tribromophenol	10.557	330	250118	103.213	ng	0.00
45) 2-Fluorobiphenyl	9.086	172	1283425	87.657	ng	0.00
79) Terphenyl-d14	12.839	244	982821	78.850	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF062422\
Data File : BF128991.D
Acq On : 24 Jun 2022 16:00
Operator : CG\JU
Sample : N3431-08
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
WC-14B-2-3

Quant Time: Jun 25 06:14:11 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF060222.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 24 00:50:48 2022
Response via : Initial Calibration

