

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF062422\
 Data File : BF128996.D
 Acq On : 24 Jun 2022 18:28
 Operator : CG\JU
 Sample : N3431-20
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 WC-14B-2-5

Quant Time: Jun 25 06:15:19 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF060222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 24 00:50:48 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.722	152	102748	20.000	ng	0.00
21) Naphthalene-d8	8.004	136	391023	20.000	ng	0.00
39) Acenaphthene-d10	9.757	164	207751	20.000	ng	-0.01
64) Phenanthrene-d10	11.245	188	335821	20.000	ng	0.00
76) Chrysene-d12	13.892	240	234551	20.000	ng	0.00
86) Perylene-d12	15.333	264	206865	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.369	112	709561	109.998	ng	0.01
7) Phenol-d6	6.387	99	824325	104.688	ng	0.00
23) Nitrobenzene-d5	7.293	82	707990	84.299	ng	-0.01
42) 2,4,6-Tribromophenol	10.557	330	269132	109.640	ng	0.00
45) 2-Fluorobiphenyl	9.087	172	1241928	83.739	ng	0.00
79) Terphenyl-d14	12.839	244	1119803	82.959	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF062422\
Data File : BF128996.D
Acq On : 24 Jun 2022 18:28
Operator : CG\JU
Sample : N3431-20
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
WC-14B-2-5

Quant Time: Jun 25 06:15:19 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF060222.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 24 00:50:48 2022
Response via : Initial Calibration

