


Data Path : Z:\HPCHEM1\BNA\_F\DATA\BF062515\  
 Data File : BF080179.D  
 Acq On : 25 Jun 2015 23:00  
 Operator : TP/IZ  
 Sample : DFTPP  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA\_F\METHODS\8270-BF061715.M  
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 Last Update : Tue Jun 23 00:53:56 2015



AutoFind: Scans 787, 788, 789; Background Corrected with Scan 785

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 51          | 198          | 10           | 80           | 45.9      | 7781    | PASS             |
| 68          | 69           | 0.00         | 2            | 0.0       | 0       | PASS             |
| 69          | 198          | 0.00         | 100          | 39.7      | 6724    | PASS             |
| 70          | 69           | 0.00         | 2            | 0.0       | 0       | PASS             |
| 127         | 198          | 10           | 80           | 48.0      | 8136    | PASS             |
| 197         | 198          | 0.00         | 2            | 0.0       | 0       | PASS             |
| 198         | 198          | 100          | 100          | 100.0     | 16949   | PASS             |
| 199         | 198          | 5            | 9            | 7.0       | 1182    | PASS             |
| 275         | 198          | 10           | 60           | 25.2      | 4270    | PASS             |
| 365         | 198          | 1            | 100          | 2.4       | 414     | PASS             |
| 441         | 442          | 0.01         | 24           | 15.7      | 2568    | PASS             |
| 442         | 198          | 50           | 100          | 96.5      | 16361   | PASS             |
| 443         | 442          | 15           | 24           | 19.6      | 3199    | PASS             |