

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF062518\  
 Data File : BF106801.D  
 Acq On : 26 Jun 2018 5:07  
 Operator : JU/SJ  
 Sample : J3648-02  
 Misc :  
 ALS Vial : 29 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 DPC-81753-FB-02

Quant Time: Jun 26 07:38:24 2018  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF061918.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Mon Jun 25 18:50:17 2018  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc  | Units | Dev(Min) |
|-----------------------------|-------|------|----------|-------|-------|----------|
| 1) 1,4-Dichlorobenzene-d4   | 6.96  | 152  | 47035    | 20.00 | ng    | 0.00     |
| 21) Naphthalene-d8          | 8.24  | 136  | 164402   | 20.00 | ng    | 0.00     |
| 38) Acenaphthene-d10        | 10.00 | 164  | 71226    | 20.00 | ng    | 0.00     |
| 63) Phenanthrene-d10        | 11.49 | 188  | 137869   | 20.00 | ng    | 0.00     |
| 75) Chrysene-d12            | 14.15 | 240  | 125306   | 20.00 | ng    | -0.02    |
| 86) Perylene-d12            | 15.69 | 264  | 87913    | 20.00 | ng    | -0.03    |
| System Monitoring Compounds |       |      |          |       |       |          |
| 5) 2-Fluorophenol           | 5.61  | 112  | 224626   | 80.87 | ng    | 0.00     |
| 7) Phenol-d6                | 6.60  | 99   | 272816   | 78.65 | ng    | -0.01    |
| 23) Nitrobenzene-d5         | 7.52  | 82   | 164466   | 55.60 | ng    | -0.01    |
| 41) 2,4,6-Tribromophenol    | 10.80 | 330  | 73300    | 88.79 | ng    | 0.00     |
| 44) 2-Fluorobiphenyl        | 9.31  | 172  | 317270   | 73.38 | ng    | -0.01    |
| 78) Terphenyl-d14           | 13.08 | 244  | 371337   | 71.78 | ng    | 0.00     |

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF062518\  
Data File : BF106801.D  
Acq On : 26 Jun 2018 5:07  
Operator : JU/SJ  
Sample : J3648-02  
Misc :  
ALS Vial : 29 Sample Multiplier: 1

Instrument :  
BNA\_F  
ClientSampleId :  
DPC-81753-FB-02

Quant Time: Jun 26 07:38:24 2018  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF061918.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Mon Jun 25 18:50:17 2018  
Response via : Initial Calibration

