

Data Path : Z:\svoasrv\HPCHEM1\BNA F\Data\BF062518\
 Data File : BF106772.D
 Acq On : 25 Jun 2018 15:38
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 26 08:27:07 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF061918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 22 11:34:03 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	64	0.00
2	1,4-Dioxane	0.607	0.495	18.5	53	0.00
3	Pyridine	1.647	1.388	15.7	55	0.00
4	n-Nitrosodimethylamine	0.699	0.575	17.7	53	0.00
5 S	2-Fluorophenol	1.181	1.042	11.8	58	0.00
6	Aniline	2.001	1.744	12.8	57	0.00
7 S	Phenol-d6	1.475	1.322	10.4	60	0.00
8	2-Chlorophenol	1.295	1.194	7.8	61	0.00
9	Benzaldehyde	1.008	0.854	15.3	56	0.00
10 C	Phenol	1.683	1.416	15.9	56	0.00
11	bis(2-Chloroethyl)ether	1.390	1.260	9.4	60	0.00
12	1,3-Dichlorobenzene	1.517	1.409	7.1	61	0.00
13 C	1,4-Dichlorobenzene	1.534	1.422	7.3	61	0.00
14	1,2-Dichlorobenzene	1.406	1.295	7.9	60	0.00
15	Benzyl Alcohol	1.184	1.039	12.2	57	0.00
16	2,2'-oxybis(1-Chloropropane	1.823	1.494	18.0	54	0.00
17	2-Methylphenol	1.109	1.080	2.6	64	0.00
18	Hexachloroethane	0.539	0.493	8.5	60	0.00
19 P	n-Nitroso-di-n-propylamine	0.972	0.825	15.1	59	0.00
20	3+4-Methylphenols	1.283	1.106	13.8	58	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	64	0.00
22	Acetophenone	0.447	0.407	8.9	60	0.00
23 S	Nitrobenzene-d5	0.360	0.318	11.7	59	0.00
24	Nitrobenzene	0.367	0.317	13.6	57	0.00
25	Isophorone	0.634	0.571	9.9	60	0.00
26 C	2-Nitrophenol	0.173	0.170	1.7	63	0.00
27	2,4-Dimethylphenol	0.268	0.244	9.0	59	0.00
28	bis(2-Chloroethoxy)methane	0.426	0.381	10.6	60	0.00
29 C	2,4-Dichlorophenol	0.281	0.267	5.0	62	0.00
30	1,2,4-Trichlorobenzene	0.309	0.310	-0.3	66	0.00
31	Naphthalene	0.935	0.850	9.1	61	0.00
32	Benzoic acid	0.180	0.167	7.2	60	0.00
33	4-Chloroaniline	0.392	0.363	7.4	62	0.00
34 C	Hexachlorobutadiene	0.201	0.208	-3.5	68	0.00
35	Caprolactam	0.091	0.087	4.4	61	0.00
36 C	4-Chloro-3-methylphenol	0.295	0.277	6.1	62	0.00
37	2-Methylnaphthalene	0.620	0.610	1.6	66	0.00
38 I	Acenaphthene-d10	1.000	1.000	0.0	69	0.00
39	1,2,4,5-Tetrachlorobenzene	0.674	0.637	5.5	68	0.00
40 P	Hexachlorocyclopentadiene	0.347	0.387	-11.5	77	0.00
41 S	2,4,6-Tribromophenol	0.232	0.231	0.4	70	0.00
42 C	2,4,6-Trichlorophenol	0.448	0.388	13.4	62	0.00
43	2,4,5-Trichlorophenol	0.467	0.407	12.8	62	0.00
44 S	2-Fluorobiphenyl	1.309	1.079	17.6	62	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA F\Data\BF062518\
 Data File : BF106772.D
 Acq On : 25 Jun 2018 15:38
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 26 08:27:07 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF061918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 22 11:34:03 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	1.594	1.464	8.2	67	0.00
46	2-Chloronaphthalene	1.255	1.070	14.7	62	0.00
47	2-Nitroaniline	0.422	0.344	18.5	57	0.00
48	Acenaphthylene	1.981	1.688	14.8	62	0.00
49	Dimethylphthalate	1.500	1.333	11.1	65	0.00
50	2,6-Dinitrotoluene	0.318	0.298	6.3	67	0.00
51 C	Acenaphthene	1.247	1.065	14.6	61	0.00
52	3-Nitroaniline	0.366	0.327	10.7	63	0.00
53 P	2,4-Dinitrophenol	0.145	0.221	-52.4#	105	0.00
54	Dibenzofuran	1.708	1.565	8.4	66	0.00
55 P	4-Nitrophenol	0.255	0.236	7.5	63	0.00
56	2,4-Dinitrotoluene	0.415	0.390	6.0	65	0.00
57	Fluorene	1.355	1.228	9.4	67	0.00
58	2,3,4,6-Tetrachlorophenol	0.371	0.352	5.1	67	0.00
59	Diethylphthalate	1.448	1.239	14.4	62	0.00
60	4-Chlorophenyl-phenylether	0.709	0.656	7.5	67	0.00
61	4-Nitroaniline	0.363	0.326	10.2	63	0.00
62	Azobenzene	1.426	1.231	13.7	63	0.00
63 I	Phenanthrene-d10	1.000	1.000	0.0	69	0.00
64	4,6-Dinitro-2-methylphenol	0.124	0.120	3.2	67	0.00
65 c	n-Nitrosodiphenylamine	0.673	0.571	15.2	62	0.00
66	4-Bromophenyl-phenylether	0.239	0.226	5.4	69	0.00
67	Hexachlorobenzene	0.250	0.243	2.8	70	0.00
68	Atrazine	0.214	0.201	6.1	67	0.00
69 C	Pentachlorophenol	0.125	0.136	-8.8	75	0.00
70	Phenanthrene	0.965	0.897	7.0	67	0.00
71	Anthracene	0.985	0.909	7.7	66	0.00
72	Carbazole	0.922	0.818	11.3	65	0.00
73	Di-n-butylphthalate	1.086	0.930	14.4	62	0.00
74 C	Fluoranthene	1.063	0.970	8.7	65	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	61	0.00
76	Benzidine	0.570	0.537	5.8	57	0.00
77	Pyrene	1.319	1.308	0.8	65	0.00
78 S	Terphenyl-d14	0.826	0.837	-1.3	64	0.00
79	Butylbenzylphthalate	0.542	0.488	10.0	57	0.00
80	Benzo(a)anthracene	1.219	1.114	8.6	59	0.00
81	3,3'-Dichlorobenzidine	0.428	0.363	15.2	54	0.00
82	Chrysene	1.121	1.009	10.0	59	0.00
83	Bis(2-ethylhexyl)phthalate	0.693	0.573	17.3	53	0.00
84 c	Di-n-octyl phthalate	1.130	0.966	14.5	55	0.00
85	Indeno(1,2,3-cd)pyrene	0.952	0.964	-1.3	69	0.00
86 I	Perylene-d12	1.000	1.000	0.0	65	0.00
87	Benzo(b)fluoranthene	1.242	1.134	8.7	61	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA F\Data\BF062518\
Data File : BF106772.D
Acq On : 25 Jun 2018 15:38
Operator : JU/SJ
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Jun 26 08:27:07 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF061918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 22 11:34:03 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	1.183	1.052	11.1	59	0.00
89 C	Benzo(a)pyrene	1.104	1.015	8.1	60	0.00
90	Dibenzo(a,h)anthracene	0.936	0.933	0.3	68	0.00
91	Benzo(a,h,i)perylene	0.915	0.952	-4.0	72	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0