

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF062624\
 Data File : BF138359.D
 Acq On : 26 Jun 2024 19:00
 Operator : CG\JU
 Sample : P3049-02
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SHOR-RAG-DRUM

Quant Time: Jun 27 01:13:15 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF061224.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 13 01:52:47 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.881	152	200198	20.000	ng	-0.01
21) Naphthalene-d8	8.163	136	310423	20.000	ng	#-0.01
39) Acenaphthene-d10	9.963	164	339533	20.000	ng	0.04
64) Phenanthrene-d10	11.404	188	495906	20.000	ng	-0.01
76) Chrysene-d12	14.039	240	406436	20.000	ng	-0.01
86) Perylene-d12	15.510	264	448317	20.000	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.522	112	1469464	122.786	ng	0.01
7) Phenol-d6	6.516	99	1292365	85.162	ng	0.00
23) Nitrobenzene-d5	7.445	82	958025	179.667	ng	-0.01
42) 2,4,6-Tribromophenol	10.727	330	392464	116.109	ng	0.01
45) 2-Fluorobiphenyl	9.257	172	1883893	88.229	ng	0.00
79) Terphenyl-d14	12.980	244	1372731	53.494	ng	-0.02

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF062624\
Data File : BF138359.D
Acq On : 26 Jun 2024 19:00
Operator : CG\JU
Sample : P3049-02
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SHOR-RAG-DRUM

Quant Time: Jun 27 01:13:15 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF061224.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jun 13 01:52:47 2024
Response via : Initial Calibration

