

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF062624\
 Data File : BF138348.D
 Acq On : 26 Jun 2024 13:12
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 26 13:41:36 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF061224.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 13 01:52:47 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	65	-0.01
2	1,4-Dioxane	40.000	35.954	10.1	57	-0.04
3	Pyridine	40.000	34.783	13.0	56	-0.04
4	n-Nitrosodimethylamine	40.000	34.080	14.8	53	-0.04
5 S	2-Fluorophenol	80.000	74.244	7.2	59	-0.01
6	Aniline	40.000	35.743	10.6	56	0.00
7 S	Phenol-d6	80.000	71.884	10.1	57	0.00
8	2-Chlorophenol	40.000	37.631	5.9	59	-0.01
9	Benzaldehyde	40.000	39.745	0.6	67	-0.01
10 C	Phenol	40.000	37.303	6.7	59	0.00
11	bis(2-Chloroethyl)ether	40.000	36.806	8.0	59	-0.01
12	1,3-Dichlorobenzene	40.000	36.584	8.5	58	-0.01
13 C	1,4-Dichlorobenzene	40.000	36.668	8.3	58	-0.01
14	1,2-Dichlorobenzene	40.000	36.798	8.0	58	0.00
15	Benzyl Alcohol	40.000	38.457	3.9	60	-0.01
16	2,2'-oxybis(1-Chloropropane	40.000	33.437	16.4	53	-0.01
17	2-Methylphenol	40.000	37.846	5.4	59	0.00
18	Hexachloroethane	40.000	38.391	4.0	60	-0.02
19 P	n-Nitroso-di-n-propylamine	40.000	35.286	11.8	57	-0.01
20	3+4-Methylphenols	40.000	36.515	8.7	57	-0.01
21 I	Naphthalene-d8	20.000	20.000	0.0	68	-0.01
22	Acetophenone	40.000	34.362	14.1	58	-0.01
23 S	Nitrobenzene-d5	80.000	73.739	7.8	60	0.00
24	Nitrobenzene	40.000	36.237	9.4	60	0.00
25	Isophorone	40.000	35.824	10.4	60	-0.01
26 C	2-Nitrophenol	40.000	46.824	-17.1	72	-0.01
27	2,4-Dimethylphenol	40.000	35.820	10.4	60	0.00
28	bis(2-Chloroethoxy)methane	40.000	36.199	9.5	61	0.00
29 C	2,4-Dichlorophenol	40.000	37.317	6.7	61	0.00
30	1,2,4-Trichlorobenzene	40.000	36.265	9.3	60	-0.01
31	Naphthalene	40.000	35.580	11.1	59	-0.01
32	Benzoic acid	40.000	41.747	-4.4	71	0.00
33	4-Chloroaniline	40.000	35.329	11.7	60	-0.01
34 C	Hexachlorobutadiene	40.000	36.959	7.6	62	-0.02
35	Caprolactam	40.000	39.595	1.0	65	0.00
36 C	4-Chloro-3-methylphenol	40.000	38.311	4.2	64	0.00
37	2-Methylnaphthalene	40.000	36.197	9.5	61	-0.01
38	1-Methylnaphthalene	40.000	36.068	9.8	61	-0.01
39 I	Acenaphthene-d10	20.000	20.000	0.0	71	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	35.638	10.9	62	-0.01
41 P	Hexachlorocyclopentadiene	40.000	37.774	5.6	61	-0.01
42 S	2,4,6-Tribromophenol	80.000	78.660	1.7	66	0.00
43 C	2,4,6-Trichlorophenol	40.000	37.958	5.1	63	-0.01
44	2,4,5-Trichlorophenol	40.000	37.666	5.8	64	0.00
45 S	2-Fluorobiphenyl	80.000	69.685	12.9	62	-0.01
46	1,1'-Biphenyl	40.000	35.528	11.2	62	-0.01
47	2-Chloronaphthalene	40.000	36.027	9.9	63	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF062624\
 Data File : BF138348.D
 Acq On : 26 Jun 2024 13:12
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 26 13:41:36 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF061224.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 13 01:52:47 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	40.001	-0.0	64	0.00
49	Acenaphthylene	40.000	36.172	9.6	63	-0.01
50	Dimethylphthalate	40.000	38.020	4.9	66	-0.01
51	2,6-Dinitrotoluene	40.000	42.165	-5.4	69	0.00
52 C	Acenaphthene	40.000	36.121	9.7	62	0.00
53	3-Nitroaniline	40.000	39.795	0.5	65	0.00
54 P	2,4-Dinitrophenol	40.000	44.901	-12.3	82	0.00
55	Dibenzofuran	40.000	35.866	10.3	62	0.00
56 P	4-Nitrophenol	40.000	36.180	9.6	59	0.00
57	2,4-Dinitrotoluene	40.000	44.473	-11.2	72	-0.01
58	Fluorene	40.000	35.166	12.1	62	-0.01
59	2,3,4,6-Tetrachlorophenol	40.000	37.610	6.0	63	-0.01
60	Diethylphthalate	40.000	38.463	3.8	66	-0.01
61	4-Chlorophenyl-phenylether	40.000	36.463	8.8	64	-0.01
62	4-Nitroaniline	40.000	38.116	4.7	63	0.00
63	Azobenzene	40.000	35.073	12.3	61	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	71	-0.01
65	4,6-Dinitro-2-methylphenol	40.000	43.134	-7.8	81	0.00
66 c	n-Nitrosodiphenylamine	40.000	35.873	10.3	63	0.00
67	4-Bromophenyl-phenylether	40.000	36.963	7.6	64	-0.01
68	Hexachlorobenzene	40.000	36.963	7.6	64	0.00
69	Atrazine	40.000	34.304	14.2	60	0.00
70 C	Pentachlorophenol	40.000	37.018	7.5	61	-0.01
71	Phenanthrene	40.000	35.492	11.3	62	-0.01
72	Anthracene	40.000	35.455	11.4	62	-0.01
73	Carbazole	40.000	34.030	14.9	59	-0.01
74	Di-n-butylphthalate	40.000	39.140	2.1	67	-0.01
75 C	Fluoranthene	40.000	33.335	16.7	58	-0.01
76 I	Chrysene-d12	20.000	20.000	0.0	62	-0.01
77	Benzydine	40.000	60.307	-50.8#	53	0.00
78	Pyrene	40.000	38.864	2.8	56	-0.01
79 S	Terphenyl-d14	80.000	79.727	0.3	59	-0.01
80	Butylbenzylphthalate	40.000	47.966	-19.9	67	-0.01
81	Benzo(a)anthracene	40.000	36.688	8.3	55	0.00
82	3,3'-Dichlorobenzidine	40.000	44.688	-11.7	67	-0.01
83	Chrysene	40.000	36.302	9.2	55	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	51.244	-28.1#	74	-0.02
85 c	Di-n-octyl phthalate	40.000	56.486	-41.2#	87	-0.01
86 I	Perylene-d12	20.000	20.000	0.0	76	-0.01
87	Indeno(1,2,3-cd)pyrene	40.000	41.362	-3.4	69	-0.02
88	Benzo(b)fluoranthene	40.000	34.363	14.1	66	-0.01
89	Benzo(k)fluoranthene	40.000	35.068	12.3	67	-0.01
90 C	Benzo(a)pyrene	40.000	36.790	8.0	68	-0.01
91	Dibenzo(a,h)anthracene	40.000	41.194	-3.0	68	-0.02
92	Benzo(g,h,i)perylene	40.000	41.638	-4.1	69	-0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF062624\
Data File : BF138348.D
Acq On : 26 Jun 2024 13:12
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Jun 26 13:41:36 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF061224.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jun 13 01:52:47 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 1