

Data Path : Z:\HPCHEM1\BNA F\DATA\BF062917\
 Data File : BF096315.D
 Acq On : 30 Jun 2017 2:06
 Operator : SJ/MA
 Sample : I3882-14
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 G54-1.5-3

Manual Integrations
 APPROVED

mohammad
 6/30/2017 2:52:24 PM

Quant Time: Jun 30 04:08:01 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF062717.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 28 15:41:49 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.77	152	141524	20.00	ng	0.00
21) Naphthalene-d8	8.05	136	534213	20.00	ng	0.00
38) Acenaphthene-d10	9.80	164	186043	20.00	ng	0.00
63) Phenanthrene-d10	11.29	188	291286	20.00	ng	0.00
75) Chrysene-d12	13.92	240	278397	20.00	ng	0.00
86) Perylene-d12	15.34	264	190033	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.38	112	242910	24.73	ng	0.00
7) Phenol-d6	6.39	99	405388	34.08	ng	0.00
23) Nitrobenzene-d5	7.33	82	227350	29.46	ng	-0.03
41) 2,4,6-Tribromophenol	10.59	330	19260	13.15	ng	0.00
44) 2-Fluorobiphenyl	9.13	172	373162	36.73	ng	-0.02
78) Terphenyl-d14	12.87	244	289704	25.97	ng	0.00
Target Compounds						Qvalue
49) Dimethylphthalate	9.52	163	68798	5.17	ng	99
74) Fluoranthene	12.49	202	73975	4.57	ng	94
77) Pyrene	12.72	202	72953	3.19	ng	97
82) Chrysene	13.94	228	36070	2.06	ng	# 94
83) Bis(2-ethylhexyl)phthalate	13.91	149	34564	3.03	ng	# 100
87) Benzo(b)fluoranthene	14.93	252	37579m	3.01	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF062917\
Data File : BF096315.D
Acq On : 30 Jun 2017 2:06
Operator : SJ/MA
Sample : I3882-14
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
G54-1.5-3

Manual Integrations
APPROVED

mohammad
6/30/2017 2:52:24 PM

Quant Time: Jun 30 04:08:01 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF062717.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 28 15:41:49 2017
Response via : Initial Calibration

