

Data Path : Z:\HPCHEM1\BNA F\DATA\BF062917\
 Data File : BF096317.D
 Acq On : 30 Jun 2017 3:00
 Operator : SJ/MA
 Sample : I3883-02 10X
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 G45-1.5-3

Manual Integrations
 APPROVED

mohammad
 6/30/2017 2:52:27 PM

Quant Time: Jun 30 04:14:18 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF062717.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 28 15:41:49 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.77	152	141853	20.00	ng	-0.01
21) Naphthalene-d8	8.05	136	525943	20.00	ng	0.00
38) Acenaphthene-d10	9.80	164	187402	20.00	ng	0.00
63) Phenanthrene-d10	11.29	188	301047	20.00	ng	0.00
75) Chrysene-d12	13.92	240	265885	20.00	ng	0.00
86) Perylene-d12	15.35	264	178394	20.00	ng	0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.37	112	100380	10.19	ng	0.00
7) Phenol-d6	6.39	99	119120	9.99	ng	0.00
23) Nitrobenzene-d5	7.32	82	55996	7.37	ng	-0.04
41) 2,4,6-Tribromophenol	10.59	330	17030	11.55	ng	0.00
44) 2-Fluorobiphenyl	9.13	172	100216	9.79	ng	-0.02
78) Terphenyl-d14	12.87	244	79476	7.46	ng	0.00
Target Compounds						
74) Fluoranthene	12.49	202	34022	2.03	ng	92
77) Pyrene	12.72	202	53113	2.43	ng	99
87) Benzo(b)fluoranthene	14.94	252	29754m	2.54	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF062917\
Data File : BF096317.D
Acq On : 30 Jun 2017 3:00
Operator : SJ/MA
Sample : I3883-02 10X
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
G45-1.5-3

Manual Integrations
APPROVED

mohammad
6/30/2017 2:52:27 PM

Quant Time: Jun 30 04:14:18 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF062717.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 28 15:41:49 2017
Response via : Initial Calibration

