

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF062918\
 Data File : BF106971.D
 Acq On : 29 Jun 2018 14:25
 Operator : JU/SJ
 Sample : J3677-11
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 147-149-B3(0-5)

Manual Integrations
 APPROVED

Sohil
 6/30/2018 10:58:00 AM

Quant Time: Jun 29 15:39:28 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF061918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 29 11:40:02 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.95	152	71215	20.00	ng	0.00
21) Naphthalene-d8	8.24	136	276885	20.00	ng	-0.01
38) Acenaphthene-d10	10.00	164	133578	20.00	ng	-0.01
63) Phenanthrene-d10	11.49	188	239187	20.00	ng	-0.01
75) Chrysene-d12	14.15	240	207817	20.00	ng	-0.02
86) Perylene-d12	15.69	264	165005	20.00	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.60	112	403432	95.93	ng	0.00
7) Phenol-d6	6.60	99	512589	97.60	ng	-0.01
23) Nitrobenzene-d5	7.52	82	319795	64.19	ng	-0.02
41) 2,4,6-Tribromophenol	10.79	330	160724	103.81	ng	-0.01
44) 2-Fluorobiphenyl	9.31	172	606242	75.10	ng	-0.02
78) Terphenyl-d14	13.07	244	617897	72.02	ng	-0.01
Target Compounds						
49) Dimethylphthalate	9.70	163	89512	8.935	ng	99
70) Phenanthrene	11.51	178	40328	3.496	ng	98
74) Fluoranthene	12.71	202	86792	6.826	ng	93
77) Pyrene	12.94	202	78015	5.693	ng	98
80) Benzo(a)anthracene	14.14	228	43525	3.436	ng	99
82) Chrysene	14.17	228	50887	4.367	ng	97
85) Indeno(1,2,3-cd)pyrene	17.27	276	22327	2.258	ng	# 88
87) Benzo(b)fluoranthene	15.24	252	55811m	5.445	ng	
88) Benzo(k)fluoranthene	15.27	252	20277m	2.077	ng	
89) Benzo(a)pyrene	15.63	252	34304	3.765	ng	95
91) Benzo(g,h,i)perylene	17.75	276	22224	2.944	ng	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF062918\
Data File : BF106971.D
Acq On : 29 Jun 2018 14:25
Operator : JU/SJ
Sample : J3677-11
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
147-149-B3(0-5)

Manual Integrations
APPROVED

Sohil
6/30/2018 10:58:00 AM

Quant Time: Jun 29 15:39:28 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF061918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 29 11:40:02 2018
Response via : Initial Calibration

