

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF062922\
 Data File : BF129038.D
 Acq On : 29 Jun 2022 16:28
 Operator : CG\JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF062922

Quant Time: Jun 29 19:49:21 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF062922.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 29 17:05:31 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	100	0.00
2	1,4-Dioxane	40.000	38.920	2.7	98	0.00
3	Pyridine	40.000	39.302	1.7	96	-0.01
4	n-Nitrosodimethylamine	40.000	39.320	1.7	98	0.00
5 S	2-Fluorophenol	80.000	77.551	3.1	98	0.00
6	Aniline	40.000	36.903	7.7	93	0.00
7 S	Phenol-d6	80.000	77.447	3.2	98	0.00
8	2-Chlorophenol	40.000	39.221	1.9	99	0.00
9	Benzaldehyde	40.000	40.654	-1.6	106	0.00
10 C	Phenol	40.000	39.116	2.2	99	0.00
11	bis(2-Chloroethyl)ether	40.000	39.142	2.1	100	0.00
12	1,3-Dichlorobenzene	40.000	38.864	2.8	99	0.00
13 C	1,4-Dichlorobenzene	40.000	39.090	2.3	100	0.00
14	1,2-Dichlorobenzene	40.000	39.061	2.3	98	0.00
15	Benzyl Alcohol	40.000	40.002	-0.0	100	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	39.062	2.3	100	0.00
17	2-Methylphenol	40.000	38.921	2.7	99	0.00
18	Hexachloroethane	40.000	39.955	0.1	102	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	39.400	1.5	100	0.00
20	3+4-Methylphenols	40.000	39.796	0.5	100	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	100	0.00
22	Acetophenone	40.000	39.610	1.0	100	0.00
23 S	Nitrobenzene-d5	80.000	82.215	-2.8	101	0.00
24	Nitrobenzene	40.000	40.596	-1.5	101	0.00
25	Isophorone	40.000	39.623	0.9	100	0.00
26 C	2-Nitrophenol	40.000	44.103	-10.3	103	0.00
27	2,4-Dimethylphenol	40.000	39.977	0.1	100	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.708	0.7	99	0.00
29 C	2,4-Dichlorophenol	40.000	40.177	-0.4	100	0.00
30	1,2,4-Trichlorobenzene	40.000	39.446	1.4	100	0.00
31	Naphthalene	40.000	39.860	0.4	99	0.00
32	Benzoic acid	40.000	42.678	-6.7	106	0.00
33	4-Chloroaniline	40.000	38.851	2.9	97	0.00
34 C	Hexachlorobutadiene	40.000	39.407	1.5	100	0.00
35	Caprolactam	40.000	40.217	-0.5	101	0.00
36 C	4-Chloro-3-methylphenol	40.000	40.069	-0.2	100	0.00
37	2-Methylnaphthalene	40.000	39.800	0.5	99	0.00
38	1-Methylnaphthalene	40.000	39.996	0.0	99	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	100	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	39.581	1.0	99	0.00
41 P	Hexachlorocyclopentadiene	40.000	41.457	-3.6	100	0.00
42 S	2,4,6-Tribromophenol	80.000	79.909	0.1	100	0.00
43 C	2,4,6-Trichlorophenol	40.000	40.733	-1.8	99	0.00
44	2,4,5-Trichlorophenol	40.000	40.316	-0.8	100	0.00
45 S	2-Fluorobiphenyl	80.000	72.258	9.7	99	0.00
46	1,1'-Biphenyl	40.000	39.862	0.3	100	0.00
47	2-Chloronaphthalene	40.000	39.383	1.5	99	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF062922\
 Data File : BF129038.D
 Acq On : 29 Jun 2022 16:28
 Operator : CG\JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF062922

Quant Time: Jun 29 19:49:21 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF062922.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 29 17:05:31 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	43.110	-7.8	103	0.00
49	Acenaphthylene	40.000	39.956	0.1	99	0.00
50	Dimethylphthalate	40.000	39.360	1.6	99	0.00
51	2,6-Dinitrotoluene	40.000	42.137	-5.3	102	0.00
52 C	Acenaphthene	40.000	39.799	0.5	99	0.00
53	3-Nitroaniline	40.000	41.926	-4.8	102	0.00
54 P	2,4-Dinitrophenol	40.000	40.333	-0.8	111	0.00
55	Dibenzofuran	40.000	39.445	1.4	99	0.00
56 P	4-Nitrophenol	40.000	42.166	-5.4	101	0.00
57	2,4-Dinitrotoluene	40.000	43.290	-8.2	102	0.00
58	Fluorene	40.000	39.480	1.3	99	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	40.317	-0.8	100	0.00
60	Diethylphthalate	40.000	39.578	1.1	100	0.00
61	4-Chlorophenyl-phenylether	40.000	39.339	1.7	99	0.00
62	4-Nitroaniline	40.000	41.583	-4.0	100	0.00
63	Azobenzene	40.000	39.739	0.7	99	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	99	0.00
65	4,6-Dinitro-2-methylphenol	40.000	43.676	-9.2	106	0.00
66 c	n-Nitrosodiphenylamine	40.000	40.097	-0.2	99	0.00
67	4-Bromophenyl-phenylether	40.000	40.182	-0.5	99	0.00
68	Hexachlorobenzene	40.000	39.404	1.5	100	0.00
69	Atrazine	40.000	40.347	-0.9	98	0.00
70 C	Pentachlorophenol	40.000	41.914	-4.8	100	0.00
71	Phenanthrene	40.000	39.225	1.9	98	0.00
72	Anthracene	40.000	39.515	1.2	98	0.00
73	Carbazole	40.000	39.541	1.1	100	0.00
74	Di-n-butylphthalate	40.000	40.055	-0.1	99	0.00
75 C	Fluoranthene	40.000	39.042	2.4	99	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	98	0.00
77	Benzydine	40.000	38.995	2.5	89	0.00
78	Pyrene	40.000	41.191	-3.0	99	0.00
79 S	Terphenyl-d14	80.000	79.068	1.2	98	0.00
80	Butylbenzylphthalate	40.000	41.569	-3.9	99	0.00
81	Benzo(a)anthracene	40.000	39.541	1.1	98	0.00
82	3,3'-Dichlorobenzidine	40.000	36.557	8.6	93	0.00
83	Chrysene	40.000	39.064	2.3	96	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	40.709	-1.8	98	0.00
85 c	Di-n-octyl phthalate	40.000	40.169	-0.4	98	0.00
86 I	Perylene-d12	20.000	20.000	0.0	96	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	39.025	2.4	97	0.00
88	Benzo(b)fluoranthene	40.000	40.278	-0.7	96	0.00
89	Benzo(k)fluoranthene	40.000	40.576	-1.4	96	0.00
90 C	Benzo(a)pyrene	40.000	39.774	0.6	95	0.00
91	Dibenzo(a,h)anthracene	40.000	39.217	2.0	97	0.00
92	Benzo(g,h,i)perylene	40.000	39.077	2.3	98	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF062922\
Data File : BF129038.D
Acq On : 29 Jun 2022 16:28
Operator : CG\JU
Sample : SSTDICV040
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
ICVBF062922

Quant Time: Jun 29 19:49:21 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF062922.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 29 17:05:31 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0