

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF062922\
 Data File : BF129038.D
 Acq On : 29 Jun 2022 16:28
 Operator : CG\JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF062922

Quant Time: Jun 29 19:49:21 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF062922.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 29 17:05:31 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	100	0.00
2	1,4-Dioxane	0.527	0.513	2.7	98	0.00
3	Pyridine	1.288	1.266	1.7	96	-0.01
4	n-Nitrosodimethylamine	0.584	0.574	1.7	98	0.00
5 S	2-Fluorophenol	1.183	1.147	3.0	98	0.00
6	Aniline	1.643	1.516	7.7	93	0.00
7 S	Phenol-d6	1.469	1.422	3.2	98	0.00
8	2-Chlorophenol	1.243	1.219	1.9	99	0.00
9	Benzaldehyde	0.774	0.688	11.1	106	0.00
10 C	Phenol	1.489	1.456	2.2	99	0.00
11	bis(2-Chloroethyl)ether	1.178	1.153	2.1	100	0.00
12	1,3-Dichlorobenzene	1.368	1.329	2.9	99	0.00
13 C	1,4-Dichlorobenzene	1.380	1.349	2.2	100	0.00
14	1,2-Dichlorobenzene	1.297	1.267	2.3	98	0.00
15	Benzyl Alcohol	1.033	1.034	-0.1	100	0.00
16	2,2'-oxybis(1-Chloropropane	1.683	1.643	2.4	100	0.00
17	2-Methylphenol	1.011	0.984	2.7	99	0.00
18	Hexachloroethane	0.484	0.484	0.0	102	0.00
19 P	n-Nitroso-di-n-propylamine	0.850	0.837	1.5	100	0.00
20	3+4-Methylphenols	1.286	1.280	0.5	100	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	100	0.00
22	Acetophenone	0.459	0.454	1.1	100	0.00
23 S	Nitrobenzene-d5	0.335	0.345	-3.0	101	0.00
24	Nitrobenzene	0.324	0.329	-1.5	101	0.00
25	Isophorone	0.567	0.561	1.1	100	0.00
26 C	2-Nitrophenol	0.151	0.166	-9.9	103	0.00
27	2,4-Dimethylphenol	0.272	0.272	0.0	100	0.00
28	bis(2-Chloroethoxy)methane	0.389	0.386	0.8	99	0.00
29 C	2,4-Dichlorophenol	0.278	0.279	-0.4	100	0.00
30	1,2,4-Trichlorobenzene	0.306	0.301	1.6	100	0.00
31	Naphthalene	0.908	0.905	0.3	99	0.00
32	Benzoic acid	0.218	0.233	-6.9	106	0.00
33	4-Chloroaniline	0.366	0.355	3.0	97	0.00
34 C	Hexachlorobutadiene	0.182	0.180	1.1	100	0.00
35	Caprolactam	0.088	0.089	-1.1	101	0.00
36 C	4-Chloro-3-methylphenol	0.282	0.283	-0.4	100	0.00
37	2-Methylnaphthalene	0.611	0.608	0.5	99	0.00
38	1-Methylnaphthalene	0.593	0.593	0.0	99	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	100	0.00
40	1,2,4,5-Tetrachlorobenzene	0.562	0.556	1.1	99	0.00
41 P	Hexachlorocyclopentadiene	0.297	0.308	-3.7	100	0.00
42 S	2,4,6-Tribromophenol	0.217	0.217	0.0	100	0.00
43 C	2,4,6-Trichlorophenol	0.383	0.390	-1.8	99	0.00
44	2,4,5-Trichlorophenol	0.407	0.411	-1.0	100	0.00
45 S	2-Fluorobiphenyl	1.250	1.129	9.7	99	0.00
46	1,1'-Biphenyl	1.438	1.433	0.3	100	0.00
47	2-Chloronaphthalene	1.110	1.093	1.5	99	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF062922\
 Data File : BF129038.D
 Acq On : 29 Jun 2022 16:28
 Operator : CG\JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF062922

Quant Time: Jun 29 19:49:21 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF062922.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 29 17:05:31 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.292	0.315	-7.9	103	0.00
49	Acenaphthylene	1.687	1.685	0.1	99	0.00
50	Dimethylphthalate	1.251	1.231	1.6	99	0.00
51	2,6-Dinitrotoluene	0.256	0.269	-5.1	102	0.00
52 C	Acenaphthene	1.099	1.093	0.5	99	0.00
53	3-Nitroaniline	0.305	0.320	-4.9	102	0.00
54 P	2,4-Dinitrophenol	0.115	0.127	-10.4	111	0.00
55	Dibenzofuran	1.600	1.578	1.4	99	0.00
56 P	4-Nitrophenol	0.249	0.263	-5.6	101	0.00
57	2,4-Dinitrotoluene	0.316	0.342	-8.2	102	0.00
58	Fluorene	1.238	1.222	1.3	99	0.00
59	2,3,4,6-Tetrachlorophenol	0.337	0.340	-0.9	100	0.00
60	Diethylphthalate	1.254	1.241	1.0	100	0.00
61	4-Chlorophenyl-phenylether	0.617	0.607	1.6	99	0.00
62	4-Nitroaniline	0.302	0.314	-4.0	100	0.00
63	Azobenzene	1.228	1.220	0.7	99	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	99	0.00
65	4,6-Dinitro-2-methylphenol	0.088	0.097	-10.2	106	0.00
66 c	n-Nitrosodiphenylamine	0.607	0.609	-0.3	99	0.00
67	4-Bromophenyl-phenylether	0.213	0.214	-0.5	99	0.00
68	Hexachlorobenzene	0.237	0.233	1.7	100	0.00
69	Atrazine	0.171	0.172	-0.6	98	0.00
70 C	Pentachlorophenol	0.141	0.148	-5.0	100	0.00
71	Phenanthrene	0.956	0.937	2.0	98	0.00
72	Anthracene	0.942	0.931	1.2	98	0.00
73	Carbazole	0.944	0.933	1.2	100	0.00
74	Di-n-butylphthalate	0.995	0.997	-0.2	99	0.00
75 C	Fluoranthene	0.999	0.975	2.4	99	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	98	0.00
77	Benzydine	0.356	0.347	2.5	89	0.00
78	Pyrene	1.386	1.428	-3.0	99	0.00
79 S	Terphenyl-d14	0.963	0.952	1.1	98	0.00
80	Butylbenzylphthalate	0.565	0.587	-3.9	99	0.00
81	Benzo(a)anthracene	1.221	1.207	1.1	98	0.00
82	3,3'-Dichlorobenzidine	0.264	0.241	8.7	93	0.00
83	Chrysene	1.134	1.107	2.4	96	0.00
84	Bis(2-ethylhexyl)phthalate	0.751	0.764	-1.7	98	0.00
85 c	Di-n-octyl phthalate	1.108	1.113	-0.5	98	0.00
86 I	Perylene-d12	1.000	1.000	0.0	96	0.00
87	Indeno(1,2,3-cd)pyrene	1.227	1.197	2.4	97	0.00
88	Benzo(b)fluoranthene	1.203	1.211	-0.7	96	0.00
89	Benzo(k)fluoranthene	1.167	1.183	-1.4	96	0.00
90 C	Benzo(a)pyrene	0.981	0.975	0.6	95	0.00
91	Dibenzo(a,h)anthracene	1.000	0.980	2.0	97	0.00
92	Benzo(g,h,i)perylene	1.003	0.980	2.3	98	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF062922\
Data File : BF129038.D
Acq On : 29 Jun 2022 16:28
Operator : CG\JU
Sample : SSTDICV040
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
ICVBF062922

Quant Time: Jun 29 19:49:21 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF062922.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 29 17:05:31 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0