

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF063015\
 Data File : BF080236.D
 Acq On : 30 Jun 2015 15:10
 Operator : TP/IZ
 Sample : PB84270BL
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB84270BL

Quant Time: Jul 01 02:14:53 2015
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF061715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 23 00:53:56 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.28	152	32954	20.00	ng	-0.05
21) Naphthalene-d8	8.57	136	131994	20.00	ng	-0.05
38) Acenaphthene-d10	10.34	164	66620	20.00	ng	-0.05
63) Phenanthrene-d10	11.85	188	141295	20.00	ng	-0.06
75) Chrysene-d12	14.87	240	147679	20.00	ng	-0.13
86) Perylene-d12	16.77	264	143092	20.00	ng	-0.16
System Monitoring Compounds						
5) 2-Fluorophenol	5.90	112	155382	83.47	ng	-0.03
7) Phenol-d6	6.89	99	191970	77.49	ng	-0.03
23) Nitrobenzene-d5	7.84	82	179366	79.11	ng	-0.03
41) 2,4,6-Tribromophenol	11.14	330	47421	72.79	ng	-0.03
44) 2-Fluorobiphenyl	9.65	172	346318	74.13	ng	-0.05
78) Terphenyl-d14	13.57	244	436598	69.05	ng	-0.10

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF063015\
Data File : BF080236.D
Acq On : 30 Jun 2015 15:10
Operator : TP/IZ
Sample : PB84270BL
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB84270BL

Quant Time: Jul 01 02:14:53 2015
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF061715.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jun 23 00:53:56 2015
Response via : Initial Calibration

