

Data Path : Z:\HPCHEM1\BNA F\DATA\BF063016\
 Data File : BF088554.D
 Acq On : 1 Jul 2016 3:15
 Operator : UM/SJ
 Sample : H3630-07
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 B1409-TP03E-1224

Quant Time: Jul 01 05:59:56 2016
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF063016.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 01 01:58:49 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.82	152	135650	20.00	ng	0.01
21) Naphthalene-d8	8.10	136	427289	20.00	ng	0.00
38) Acenaphthene-d10	9.86	164	174983	20.00	ng	0.01
63) Phenanthrene-d10	11.34	188	303113	20.00	ng	0.01
75) Chrysene-d12	13.98	240	195015	20.00	ng	0.02
86) Perylene-d12	15.41	264	174943	20.00	ng	0.06
System Monitoring Compounds						
5) 2-Fluorophenol	5.42	112	1049525	115.95	ng	0.01
7) Phenol-d6	6.46	99	1382970	118.25	ng	0.01
23) Nitrobenzene-d5	7.38	82	808486	99.16	ng	0.00
41) 2,4,6-Tribromophenol	10.64	330	197055	121.27	ng	0.00
44) 2-Fluorobiphenyl	9.19	172	1186982	107.15	ng	0.01
78) Terphenyl-d14	12.93	244	687905	78.57	ng	0.01
Target Compounds						
13) 1,4-Dichlorobenzene	6.83	146	26569	2.50	ng	# 90
49) Dimethylphthalate	9.58	163	218293	17.51	ng	# 94
77) Pyrene	12.78	202	47905	2.90	ng	# 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF063016\
Data File : BF088554.D
Acq On : 1 Jul 2016 3:15
Operator : UM/SJ
Sample : H3630-07
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
B1409-TP03E-1224

Quant Time: Jul 01 05:59:56 2016
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF063016.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jul 01 01:58:49 2016
Response via : Initial Calibration

