

Data Path : Z:\HPCHEM1\BNA F\DATA\BF063017\
 Data File : BF096361.D
 Acq On : 1 Jul 2017 2:52
 Operator : SJ/MA
 Sample : I3930-10
 Misc :
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :

Quant Time: Jul 01 04:05:55 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF062717.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 28 15:41:49 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.76	152	145587	20.00	ng	-0.02
21) Naphthalene-d8	8.04	136	543718	20.00	ng	-0.02
38) Acenaphthene-d10	9.79	164	198956	20.00	ng	-0.02
63) Phenanthrene-d10	11.27	188	314522	20.00	ng	-0.01
75) Chrysene-d12	13.90	240	264341	20.00	ng	-0.01
86) Perylene-d12	15.33	264	171822	20.00	ng	-0.01

System Monitoring Compounds						
5) 2-Fluorophenol	5.37	112	774042	76.60	ng	-0.01
7) Phenol-d6	6.39	99	925991	75.68	ng	0.00
23) Nitrobenzene-d5	7.32	82	505347	64.34	ng	-0.04
41) 2,4,6-Tribromophenol	10.57	330	156549	99.98	ng	-0.02
44) 2-Fluorobiphenyl	9.12	172	762793	70.21	ng	-0.03
78) Terphenyl-d14	12.86	244	580597	54.81	ng	-0.01


Target Compounds				Qvalue		
9) Benzaldehyde	6.30	77	5290	Below Cal	#	74
19) n-Nitroso-di-n-propylamine	7.32	70	66479	9.07	ng	# 87
25) Isophorone	7.32	82	502401	28.17	ng	# 67
31) Naphthalene	8.06	128	65050	2.48	ng	100
32) Benzoic acid	7.70	122	680	8.88	ng	91
37) 2-Methylnaphthalene	8.75	142	52597	3.09	ng	97
49) Dimethylphthalate	9.51	163	115889	8.14	ng	98
53) 2,4-Dinitrophenol	10.07	184	251	7.80	ng	# 68
57) Fluorene	10.34	166	5461	Below Cal	#	93
60) 4-Chlorophenyl-phenylether	10.07	204	309	Below Cal	#	11
64) 4,6-Dinitro-2-methylphenol	10.57	198	342	8.92	ng	# 1
66) 4-Bromophenyl-phenylether	10.57	248	8427	2.60	ng	# 1
70) Phenanthrene	11.29	178	108101	Below Cal		97
74) Fluoranthene	12.48	202	155346	8.89	ng	96
77) Pyrene	12.71	202	140343	6.46	ng	98
80) Benzo(a)anthracene	13.89	228	68601	4.26	ng	93
82) Chrysene	13.93	228	68828	4.13	ng	# 93
85) Indeno(1,2,3-cd)pyrene	16.71	276	33113	2.39	ng	# 92
87) Benzo(b)fluoranthene	14.92	252	94899	8.41	ng	# 82
88) Benzo(k)fluoranthene	14.92	252	94899	9.21	ng	# 86
89) Benzo(a)pyrene	15.26	252	36570	3.67	ng	# 77
91) Benzo(g,h,i)perylene	17.13	276	30308	3.74	ng	# 81


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF063017\
Data File : BF096361.D
Acq On : 1 Jul 2017 2:52
Operator : SJ/MA
Sample : I3930-10
Misc :
ALS Vial : 38 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :

Quant Time: Jul 01 04:05:55 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF062717.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 28 15:41:49 2017
Response via : Initial Calibration






#1




1,4-Dichlorobenzene-d4
Concen: 20.00 ng
RT: 6.76 min Scan# 794
Delta R.T. -0.02 min
Lab File: BF096361.D
Acq: 1 Jul 2017 2:52

Instrument :
BNA_F
ClientSampleId :

Tot Ion:152 Resp: 145587
Ion Ratio Lower Upper
152 100
150 157.5 126.2 189.2
115 65.6 52.2 78.2




Abundance Ion 152.00 (151.70 to 152.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 115.00 (114.70 to 115.70): E





#5


2-Fluorophenol
Concen: 76.60 ng
RT: 5.37 min Scan# 558
Delta R.T. -0.01 min
Lab File: BF096361.D
Acq: 1 Jul 2017 2:52

Tgt Ion:112 Resp: 774042
Ion Ratio Lower Upper
112 100
64 67.4 46.0 69.0
63 32.6 23.4 35.0



Abundance Ion 112.00 (111.70 to 112.70): E
Ion 64.00 (63.70 to 64.70): BF0
Ion 63.00 (62.70 to 63.70): BF0





#7

Phenol-d6

Concen: 75.68 ng

RT: 6.39 min Scan# 732

Delta R.T. -0.00 min

Lab File: BF096361.D

Acq: 1 Jul 2017 2:52

Instrument :

BNA_F

ClientSampleId :





Tot Ion: 99 Resp: 925991

Ion Ratio Lower Upper

99 100

42 20.3 13.5 20.3#

71 28.4 24.5 36.7



#19

n-Nitroso-di-n-propylamine

Concen: 9.07 ng

RT: 7.32 min Scan# 889

Delta R.T. 0.14 min

Lab File: BF096361.D

Acq: 1 Jul 2017 2:52

Tgt Ion: 70 Resp: 66479




Ion Ratio Lower Upper


70 100

42 57.4 47.2 70.8

101 0.0 7.2 10.8#

130 2.4 15.9 23.9#









#21
Naphthalene-d8
Concen: 20.00 ng
RT: 8.04 min Scan# 1012
Delta R.T. -0.02 min
Lab File: BF096361.D
Acq: 1 Jul 2017 2:52

Instrument :
BNA_F
ClientSampleId :

Tot Ion:136	Resp:	543718
Ion Ratio	Lower	Upper
136	100	
137	10.9	8.5 12.7
54	11.0	4.9 7.3#
68	6.9	4.1 6.1#




Abundance Ion 136.00 (135.70 to 136.70): E
Ion 137.00 (136.70 to 137.70): E
Ion 54.00 (53.70 to 54.70): BF0
Ion 68.00 (67.70 to 68.70): BF0






#23
Nitrobenzene-d5
Concen: 64.34 ng
RT: 7.32 min Scan# 889
Delta R.T. -0.04 min
Lab File: BF096361.D
Acq: 1 Jul 2017 2:52

Tgt Ion: 82	Resp:	505347
Ion Ratio	Lower	Upper
82	100	
128	47.1	34.0 51.0
54	63.9	42.2 63.2#



Abundance Ion 82.00 (81.70 to 82.70): BF0
Ion 128.00 (127.70 to 128.70): E
Ion 54.00 (53.70 to 54.70): BF0





#25

Isophorone

Concen: 28.17 ng

RT: 7.32 min Scan# 889

Delta R.T. -0.28 min

Lab File: BF096361.D

Acq: 1 Jul 2017 2:52

Instrument :

BNA_F

ClientSampleId :





Tot Ion: 82 Resp: 502401

Ion Ratio Lower Upper

82 100

95 0.1 5.3 7.9#

138 0.0 13.1 19.7#



#31

Naphthalene

Concen: 2.48 ng

RT: 8.06 min Scan# 1016

Delta R.T. -0.02 min

Lab File: BF096361.D

Acq: 1 Jul 2017 2:52




Tgt Ion: 128 Resp: 65050


Ion Ratio Lower Upper

128 100

129 11.4 9.0 13.6

127 13.5 10.8 16.2





#32

Benzoic acid

Concen: 8.88 ng

RT: 7.70 min Scan# 954

Delta R.T. -0.19 min

Lab File: BF096361.D

Acq: 1 Jul 2017 2:52

Instrument :

BNA_F

ClientSampleId :





Tot Ion:122 Resp: 680

Ion Ratio Lower Upper

122 100

105 131.9 103.3 143.3

77 103.9 73.7 113.7



#37

2-Methylnaphthalene

Concen: 3.09 ng

RT: 8.75 min Scan# 1133

Delta R.T. -0.02 min

Lab File: BF096361.D

Acq: 1 Jul 2017 2:52




Tgt Ion:142 Resp: 52597


Ion Ratio Lower Upper

142 100

141 87.7 71.3 106.9

115 29.2 27.1 40.7









#38
 Acenaphthene-d10
 Concen: 20.00 ng
 RT: 9.79 min Scan# 1310
 Delta R.T. -0.02 min
 Lab File: BF096361.D
 Acq: 1 Jul 2017 2:52

Instrument :
 BNA_F
 ClientSampleId :

Tot Ion	Ratio	Lower	Upper
164	100		
162	102.8	82.6	123.8
160	44.2	36.2	54.2




Abundance Ion 164.00 (163.70 to 164.70): E
 Ion 162.00 (161.70 to 162.70): E
 Ion 160.00 (159.70 to 160.70): E






#41
 2,4,6-Tribromophenol
 Concen: 99.98 ng
 RT: 10.57 min Scan# 1443
 Delta R.T. -0.02 min
 Lab File: BF096361.D
 Acq: 1 Jul 2017 2:52

Tgt Ion	Ratio	Lower	Upper
330	100		
332	95.9	76.6	115.0
141	29.8	31.4	47.2



Abundance Ion 329.80 (329.50 to 330.50): E
 Ion 331.80 (331.50 to 332.50): E
 Ion 141.00 (140.70 to 141.70): E









#44
2-Fluorobiphenyl
Concen: 70.21 ng
RT: 9.12 min Scan# 1195
Delta R.T. -0.03 min
Lab File: BF096361.D
Acq: 1 Jul 2017 2:52

Instrument :
BNA_F
ClientSampled :

Tot Ion:172 Resp: 762793
Ion Ratio Lower Upper
172 100
171 35.3 29.6 44.4
170 24.0 19.8 29.8




Abundance Ion 172.00 (171.70 to 172.70): E
Ion 171.00 (170.70 to 171.70): E
Ion 170.00 (169.70 to 170.70): E






#49
Dimethylphthalate
Concen: 8.14 ng
RT: 9.51 min Scan# 1262
Delta R.T. -0.02 min
Lab File: BF096361.D
Acq: 1 Jul 2017 2:52

Tgt Ion:163 Resp: 115889
Ion Ratio Lower Upper
163 100
194 3.7 2.6 4.0
164 9.7 8.4 12.6



Abundance Ion 163.00 (162.70 to 163.70): E
Ion 194.00 (193.70 to 194.70): E
Ion 164.00 (163.70 to 164.70): E









#53
2,4-Dinitrophenol
Concen: 7.80 ng
RT: 10.07 min Scan# 1358
Delta R.T. 0.20 min
Lab File: BF096361.D
Acq: 1 Jul 2017 2:52

Instrument :
BNA_F
ClientSampleId :

Tot Ion:184 Resp: 251
Ion Ratio Lower Upper
184 100
63 126.9 62.7 94.1#
154 87.5 81.1 121.7




Abundance Ion 184.00 (183.70 to 184.70): E
Ion 63.00 (62.70 to 63.70): BF0
Ion 154.00 (153.70 to 154.70): E






#57
Fluorene
Concen: Below Cal
RT: 10.34 min Scan# 1403
Delta R.T. -0.03 min
Lab File: BF096361.D
Acq: 1 Jul 2017 2:52

Tgt Ion:166 Resp: 5461
Ion Ratio Lower Upper
166 100
165 96.1 78.8 118.2
167 29.1 10.6 16.0#



Abundance Ion 166.00 (165.70 to 166.70): E
Ion 165.00 (164.70 to 165.70): E
Ion 167.00 (166.70 to 167.70): E









#63
Phenanthrene-d10
Concen: 20.00 ng
RT: 11.27 min Scan# 1562
Delta R.T. -0.01 min
Lab File: BF096361.D
Acq: 1 Jul 2017 2:52




Instrument :
BNA_F
ClientSampleId :


Tot Ion:188 Resp: 314522
Ion Ratio Lower Upper
188 100
94 11.9 9.4 14.2
80 12.6 10.2 15.2



#64
4,6-Dinitro-2-methylphenol
Concen: 8.92 ng
RT: 10.57 min Scan# 1443
Delta R.T. 0.16 min
Lab File: BF096361.D
Acq: 1 Jul 2017 2:52

Tgt Ion:198 Resp: 342
Ion Ratio Lower Upper
198 100
51 601.4 53.2 93.2#
105 341.2 45.8 85.8#









#66
 4-Bromophenyl-phenylether
 Concen: 2.60 ng
 RT: 10.57 min Scan# 1443
 Delta R.T. -0.26 min
 Lab File: BF096361.D
 Acq: 1 Jul 2017 2:52

Instrument :
 BNA_F
 ClientSampleId :

Tot Ion:248 Resp: 8427
 Ion Ratio Lower Upper
 248 100
 250 193.6 76.8 115.2#
 141 592.4 68.6 103.0#




Abundance Ion 248.00 (247.70 to 248.70): E
 Ion 250.00 (249.70 to 250.70): E
 Ion 141.00 (140.70 to 141.70): E








#74
 Fluoranthene
 Concen: 8.89 ng
 RT: 12.48 min Scan# 1767
 Delta R.T. -0.01 min
 Lab File: BF096361.D
 Acq: 1 Jul 2017 2:52

Tgt Ion:202 Resp: 155346
 Ion Ratio Lower Upper
 202 100
 101 16.9 0.0 33.2
 203 17.9 0.0 38.1



Abundance Ion 202.00 (201.70 to 202.70): E
 Ion 101.00 (100.70 to 101.70): E
 Ion 203.00 (202.70 to 203.70): E





#75

Chrysene-d12

Concen: 20.00 ng

RT: 13.90 min Scan# 2009

Delta R.T. -0.01 min

Lab File: BF096361.D

Acq: 1 Jul 2017 2:52

Instrument :

BNA_F

ClientSampleId :





Tot Ion:240 Resp: 264341

Ion Ratio Lower Upper

240 100

120 15.0 5.8 8.8#

236 23.7 20.5 30.7



#77

Pyrene

Concen: 6.46 ng

RT: 12.71 min Scan# 1806

Delta R.T. -0.01 min

Lab File: BF096361.D

Acq: 1 Jul 2017 2:52


Tgt Ion:202 Resp: 140343


Ion Ratio Lower Upper

202 100

200 21.2 17.6 26.4

203 17.2 14.4 21.6





#78

Terphenyl-d14

Concen: 54.81 ng

RT: 12.86 min Scan# 1831

Delta R.T. -0.01 min

Lab File: BF096361.D

Acq: 1 Jul 2017 2:52

Instrument :

BNA_F

ClientSampleId :


Tot Ion:244 Resp: 580597

Ion Ratio Lower Upper

244 100

212 6.3 6.0 9.0




122 14.6 10.3 15.5



Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E



#80

Benzo(a)anthracene

Concen: 4.26 ng

RT: 13.89 min Scan# 2007

Delta R.T. -0.01 min

Lab File: BF096361.D

Acq: 1 Jul 2017 2:52


Tgt Ion:228 Resp: 68601

Ion Ratio Lower Upper

228 100

226 31.4 22.6 33.8



229 23.9 16.3 24.5




Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E





#82

Chrysene

Concen: 4.13 ng

RT: 13.93 min Scan# 2013

Delta R.T. -0.01 min

Lab File: BF096361.D

Acq: 1 Jul 2017 2:52

Instrument :

BNA_F

ClientSampled :


Tot Ion:228 Resp: 68828

Ion Ratio Lower Upper

228 100

226 33.1 24.9 37.3




229 24.9 15.8 23.6#



Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E



#85

Indeno(1,2,3-cd)pyrene

Concen: 2.39 ng

RT: 16.71 min Scan# 2486

Delta R.T. -0.01 min

Lab File: BF096361.D

Acq: 1 Jul 2017 2:52


Tgt Ion:276 Resp: 33113

Ion Ratio Lower Upper

276 100

138 31.3 27.8 41.6



277 31.1 20.2 30.4#




Abundance Ion 276.00 (275.70 to 276.70): E

Ion 138.00 (137.70 to 138.70): E

Ion 277.00 (276.70 to 277.70): E









#86
Pervlene-d12
Concen: 20.00 ng
RT: 15.33 min Scan# 2251
Delta R.T. -0.01 min
Lab File: BF096361.D
Acq: 1 Jul 2017 2:52

Instrument :
BNA_F
ClientSampleId :

Tot Ion	Ratio	Lower	Upper
264	100		
260	24.0	19.5	29.3
265	23.2	17.2	25.8




Abundance Ion 264.00 (263.70 to 264.70): E
Ion 260.00 (259.70 to 260.70): E
Ion 265.00 (264.70 to 265.70): E






#87
Benzo(b)fluoranthene
Concen: 8.41 ng
RT: 14.92 min Scan# 2182
Delta R.T. -0.04 min
Lab File: BF096361.D
Acq: 1 Jul 2017 2:52

Tgt Ion	Ratio	Lower	Upper
252	100		
253	27.3	18.1	27.1#
125	25.6	9.8	14.8#



Abundance Ion 252.00 (251.70 to 252.70): E
Ion 253.00 (252.70 to 253.70): E
Ion 125.00 (124.70 to 125.70): E





#88

Benzo(k)fluoranthene

Concen: 9.21 ng

RT: 14.92 min Scan# 2182

Delta R.T. -0.04 min

Lab File: BF096361.D





Acq: 1 Jul 2017 2:52

Instrument :

BNA_F

ClientSampleId :

Tot Ion:252	Resp:	94899
Ion Ratio	Lower	Upper
252	100	
253	27.3	18.4 27.6
125	25.6	13.1 19.7#



#89

Benzo(a)pyrene

Concen: 3.67 ng




RT: 15.26 min Scan# 2240


Delta R.T. -0.01 min

Lab File: BF096361.D

Acq: 1 Jul 2017 2:52

Tgt Ion:252	Resp:	36570
Ion Ratio	Lower	Upper
252	100	
253	26.9	17.5 26.3#
125	31.4	11.0 16.4#





#91

Benzo(a,h,i)perylene

Concen: 3.74 ng

RT: 17.13 min Scan# 2557

Delta R.T. -0.01 min

Lab File: BF096361.D

Acq: 1 Jul 2017 2:52

Instrument :

BNA_F

ClientSampleId :


Tot Ion:276 Resp: 30308

Ion Ratio Lower Upper

276 100

277 34.4 18.6 28.0#

138 40.2 25.4 38.0#



Abundance Ion 276.00 (275.70 to 276.70): E

Ion 277.00 (276.70 to 277.70): E

Ion 138.00 (137.70 to 138.70): E

