




Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF063021\
Data File : BF124491.D
Acq On : 30 Jun 2021 14:00
Operator : JU/CG
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jun 30 14:33:15 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF062321.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jun 29 12:49:33 2021
Response via : Initial Calibration



TIC: BF124491.D\data.ms

(77) Benzidine




12.557min (+ 0.006) 0.00 ng

response 295972

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.10	14.41
183.00	9.80	13.83#
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF063021\
Data File : BF124491.D
Acq On : 30 Jun 2021 14:00
Operator : JU/CG
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jun 30 14:33:15 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF062321.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jun 29 12:49:33 2021
Response via : Initial Calibration



TIC: BF124491.D\data.ms

(70) Pentachlorophenol (C)

11.027min (-0.006) 0.00 ng

response 54683

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	64.50	58.82
264.00	65.70	58.54
0.00	0.00	0.00