

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF063022\
 Data File : BF129048.D
 Acq On : 30 Jun 2022 09:56
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jul 01 05:40:22 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF062922.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 29 17:05:31 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	100	0.00
2	1,4-Dioxane	40.000	38.981	2.5	98	0.00
3	Pyridine	40.000	40.397	-1.0	99	-0.02
4	n-Nitrosodimethylamine	40.000	38.182	4.5	95	0.00
5 S	2-Fluorophenol	80.000	78.352	2.1	99	0.00
6	Aniline	40.000	38.913	2.7	98	0.00
7 S	Phenol-d6	80.000	77.824	2.7	98	0.00
8	2-Chlorophenol	40.000	39.378	1.6	100	0.00
9	Benzaldehyde	40.000	41.153	-2.9	107	0.00
10 C	Phenol	40.000	38.904	2.7	99	0.00
11	bis(2-Chloroethyl)ether	40.000	39.205	2.0	100	0.00
12	1,3-Dichlorobenzene	40.000	39.632	0.9	101	0.00
13 C	1,4-Dichlorobenzene	40.000	39.732	0.7	101	0.00
14	1,2-Dichlorobenzene	40.000	39.346	1.6	99	0.00
15	Benzyl Alcohol	40.000	38.699	3.3	96	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	38.852	2.9	99	0.00
17	2-Methylphenol	40.000	39.249	1.9	100	0.00
18	Hexachloroethane	40.000	39.723	0.7	101	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	38.986	2.5	99	0.00
20	3+4-Methylphenols	40.000	40.214	-0.5	101	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	102	0.00
22	Acetophenone	40.000	38.808	3.0	100	0.00
23 S	Nitrobenzene-d5	80.000	81.097	-1.4	101	0.00
24	Nitrobenzene	40.000	40.368	-0.9	102	0.00
25	Isophorone	40.000	38.805	3.0	100	0.00
26 C	2-Nitrophenol	40.000	44.348	-10.9	106	0.00
27	2,4-Dimethylphenol	40.000	39.957	0.1	101	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.288	1.8	100	0.00
29 C	2,4-Dichlorophenol	40.000	39.964	0.1	101	0.00
30	1,2,4-Trichlorobenzene	40.000	39.376	1.6	101	0.00
31	Naphthalene	40.000	39.693	0.8	100	0.00
32	Benzoic acid	40.000	40.246	-0.6	102	0.00
33	4-Chloroaniline	40.000	39.578	1.1	101	0.00
34 C	Hexachlorobutadiene	40.000	39.539	1.2	102	0.00
35	Caprolactam	40.000	39.269	1.8	100	0.00
36 C	4-Chloro-3-methylphenol	40.000	39.696	0.8	100	0.00
37	2-Methylnaphthalene	40.000	39.738	0.7	101	0.00
38	1-Methylnaphthalene	40.000	39.611	1.0	100	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	101	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	39.844	0.4	101	0.00
41 P	Hexachlorocyclopentadiene	40.000	42.152	-5.4	103	0.00
42 S	2,4,6-Tribromophenol	80.000	81.182	-1.5	103	0.00
43 C	2,4,6-Trichlorophenol	40.000	40.501	-1.3	100	0.00
44	2,4,5-Trichlorophenol	40.000	40.099	-0.2	101	0.00
45 S	2-Fluorobiphenyl	80.000	72.250	9.7	100	0.00
46	1,1'-Biphenyl	40.000	39.803	0.5	102	0.00
47	2-Chloronaphthalene	40.000	39.499	1.3	101	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF063022\
 Data File : BF129048.D
 Acq On : 30 Jun 2022 09:56
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jul 01 05:40:22 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF062922.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 29 17:05:31 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	42.985	-7.5	104	0.00
49	Acenaphthylene	40.000	40.149	-0.4	101	0.00
50	Dimethylphthalate	40.000	39.759	0.6	102	0.00
51	2,6-Dinitrotoluene	40.000	42.525	-6.3	104	0.00
52 C	Acenaphthene	40.000	40.264	-0.7	102	0.00
53	3-Nitroaniline	40.000	41.794	-4.5	103	0.00
54 P	2,4-Dinitrophenol	40.000	41.111	-2.8	115	0.00
55	Dibenzofuran	40.000	39.606	1.0	100	0.00
56 P	4-Nitrophenol	40.000	40.081	-0.2	97	0.00
57	2,4-Dinitrotoluene	40.000	43.998	-10.0	105	0.00
58	Fluorene	40.000	39.429	1.4	101	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	41.028	-2.6	103	0.00
60	Diethylphthalate	40.000	39.484	1.3	101	0.00
61	4-Chlorophenyl-phenylether	40.000	40.063	-0.2	102	0.00
62	4-Nitroaniline	40.000	41.718	-4.3	102	0.00
63	Azobenzene	40.000	39.534	1.2	100	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	101	0.00
65	4,6-Dinitro-2-methylphenol	40.000	45.967	-14.9	113	0.00
66 c	n-Nitrosodiphenylamine	40.000	40.010	-0.0	101	0.00
67	4-Bromophenyl-phenylether	40.000	40.497	-1.2	101	0.00
68	Hexachlorobenzene	40.000	39.696	0.8	102	0.00
69	Atrazine	40.000	41.821	-4.6	104	0.00
70 C	Pentachlorophenol	40.000	41.621	-4.1	100	0.00
71	Phenanthrene	40.000	39.604	1.0	100	0.00
72	Anthracene	40.000	39.537	1.2	100	0.00
73	Carbazole	40.000	39.503	1.2	101	0.00
74	Di-n-butylphthalate	40.000	40.241	-0.6	101	0.00
75 C	Fluoranthene	40.000	39.348	1.6	101	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	104	0.00
77	Benzydine	40.000	60.605	-51.5#	148	0.00
78	Pyrene	40.000	39.556	1.1	102	0.00
79 S	Terphenyl-d14	80.000	77.304	3.4	103	0.00
80	Butylbenzylphthalate	40.000	41.072	-2.7	105	0.00
81	Benzo(a)anthracene	40.000	39.011	2.5	103	0.00
82	3,3'-Dichlorobenzidine	40.000	41.196	-3.0	112	0.00
83	Chrysene	40.000	39.195	2.0	103	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	40.803	-2.0	105	0.00
85 c	Di-n-octyl phthalate	40.000	40.611	-1.5	106	0.00
86 I	Perylene-d12	20.000	20.000	0.0	108	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	39.539	1.2	112	0.00
88	Benzo(b)fluoranthene	40.000	40.814	-2.0	110	0.00
89	Benzo(k)fluoranthene	40.000	37.634	5.9	101	0.00
90 C	Benzo(a)pyrene	40.000	39.173	2.1	106	0.00
91	Dibenzo(a,h)anthracene	40.000	40.074	-0.2	112	0.00
92	Benzo(g,h,i)perylene	40.000	39.194	2.0	111	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF063022\
Data File : BF129048.D
Acq On : 30 Jun 2022 09:56
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Jul 01 05:40:22 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF062922.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 29 17:05:31 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0