

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF063022\
 Data File : BF129069.D
 Acq On : 30 Jun 2022 20:30
 Operator : CG\JU
 Sample : N3434-06
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 B2205-35

Quant Time: Jul 01 05:44:26 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF062922.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 29 17:05:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.957	152	410473	20.000	ng	0.00
21) Naphthalene-d8	8.239	136	1552360	20.000	ng	0.00
39) Acenaphthene-d10	9.998	164	809663	20.000	ng	0.00
64) Phenanthrene-d10	11.492	188	1446399	20.000	ng	0.00
76) Chrysene-d12	14.139	240	1024497	20.000	ng	-0.01
86) Perylene-d12	15.663	264	792638	20.000	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.593	112	3075792	126.684	ng	0.00
7) Phenol-d6	6.581	99	3829121	126.985	ng	0.00
23) Nitrobenzene-d5	7.522	82	2399349	92.190	ng	0.00
42) 2,4,6-Tribromophenol	10.792	330	1186320	134.750	ng	0.00
45) 2-Fluorobiphenyl	9.316	172	4081150	80.638	ng	0.00
79) Terphenyl-d14	13.086	244	4285331	86.863	ng	0.00


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF063022\
Data File : BF129069.D
Acq On : 30 Jun 2022 20:30
Operator : CG\JU
Sample : N3434-06
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
B2205-35

Quant Time: Jul 01 05:44:26 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF062922.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 29 17:05:31 2022
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.957 min Scan# 795
Delta R.T. -0.006 min
Lab File: BF129069.D
Acq: 30 Jun 2022 20:30




Instrument :
BNA_F
ClientSampleId :
B2205-35


Tgt Ion:152 Resp: 410473
Ion Ratio Lower Upper
152 100
150 154.2 123.7 185.5
115 56.9 49.9 74.9



#5
2-Fluorophenol
Concen: 126.684 ng
RT: 5.593 min Scan# 562
Delta R.T. 0.006 min
Lab File: BF129069.D
Acq: 30 Jun 2022 20:30

Tgt Ion:112 Resp: 3075792
Ion Ratio Lower Upper
112 100
64 59.9 53.3 79.9
63 30.3 28.0 42.0









#7
Phenol-d6
Concen: 126.985 ng
RT: 6.581 min Scan# 71
Delta R.T. -0.006 min
Lab File: BF129069.D
Acq: 30 Jun 2022 20:30




Instrument :
BNA_F
ClientSampleId :
B2205-35


Tgt Ion: 99 Resp: 3829121
Ion Ratio Lower Upper
99 100
42 17.9 17.3 25.9
71 35.0 28.9 43.3



#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.239 min Scan# 1012
Delta R.T. -0.006 min
Lab File: BF129069.D
Acq: 30 Jun 2022 20:30

Tgt Ion: 136 Resp: 1552360
Ion Ratio Lower Upper
136 100
137 11.3 8.8 13.2
54 7.8 7.5 11.3
68 6.3 4.7 7.1





#23

Nitrobenzene-d5

Concen: 92.190 ng

RT: 7.522 min Scan# 891

Delta R.T. -0.006 min

Lab File: BF129069.D

Acq: 30 Jun 2022 20:30

Instrument :





BNA_F

ClientSampleId :

B2205-35

Tgt Ion: 82 Resp: 2399349

Ion	Ratio	Lower	Upper
82	100		
128	49.5	34.0	51.0
54	49.5	46.9	70.3



#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 9.998 min Scan# 1311




Delta R.T. -0.006 min


Lab File: BF129069.D

Acq: 30 Jun 2022 20:30

Tgt Ion: 164 Resp: 809663

Ion	Ratio	Lower	Upper
164	100		
162	100.0	79.5	119.3
160	46.0	34.2	51.4





#42

2,4,6-Tribromophenol

Concen: 134.750 ng

RT: 10.792 min Scan# 1447

Delta R.T. -0.006 min

Lab File: BF129069.D

Acq: 30 Jun 2022 20:30

Instrument :

BNA_F

ClientSampleId :

B2205-35





Tgt Ion:330 Resp: 1186320

Ion Ratio Lower Upper

330 100

332 95.7 76.8 115.2

141 29.5 28.0 42.0



#45

2-Fluorobiphenyl

Concen: 80.638 ng

RT: 9.316 min Scan# 1195

Delta R.T. -0.006 min

Lab File: BF129069.D

Acq: 30 Jun 2022 20:30




Tgt Ion:172 Resp: 4081150


Ion Ratio Lower Upper

172 100

171 40.2 28.8 43.2

170 27.6 19.8 29.8









#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.492 min Scan# 11
Delta R.T. -0.006 min
Lab File: BF129069.D
Acq: 30 Jun 2022 20:30




Instrument :
BNA_F
ClientSampleId :
B2205-35


Tgt Ion:188 Resp: 1446399
Ion Ratio Lower Upper
188 100
94 9.5 7.0 10.6
80 10.0 7.6 11.4



#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.139 min Scan# 2015
Delta R.T. -0.012 min
Lab File: BF129069.D
Acq: 30 Jun 2022 20:30

Tgt Ion:240 Resp: 1024497
Ion Ratio Lower Upper
240 100
120 11.6 8.9 13.3
236 25.6 20.8 31.2









#79
 Terphenyl-d14
 Concen: 86.863 ng
 RT: 13.086 min Scan# 1836
 Delta R.T. -0.000 min
 Lab File: BF129069.D
 Acq: 30 Jun 2022 20:30

Instrument :
 BNA_F
 ClientSampleId :
 B2205-35

Tgt Ion:244 Resp: 4285331
 Ion Ratio Lower Upper
 244 100
 212 8.1 6.1 9.1
 122 11.9 8.8 13.2



#86
 Perylene-d12
 Concen: 20.000 ng
 RT: 15.663 min Scan# 2274
 Delta R.T. -0.012 min
 Lab File: BF129069.D
 Acq: 30 Jun 2022 20:30

Tgt Ion:264 Resp: 792638
 Ion Ratio Lower Upper
 264 100
 260 24.0 18.4 27.6
 265 21.8 17.2 25.8

