

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF070116\
 Data File : BF088579.D
 Acq On : 1 Jul 2016 17:17
 Operator : UM/SJ
 Sample : H3808-21 5X
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SB-11-COMP

Manual Integrations
 APPROVED

sohil
 7/5/2016 7:31:33 PM

Quant Time: Jul 01 18:03:46 2016
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF063016.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 30 18:01:55 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.81	152	138175	20.00	ng	0.00
21) Naphthalene-d8	8.09	136	531379	20.00	ng	-0.01
38) Acenaphthene-d10	9.84	164	228472	20.00	ng	-0.01
63) Phenanthrene-d10	11.33	188	424807	20.00	ng	0.00
75) Chrysene-d12	13.95	240	309683	20.00	ng	0.00
86) Perylene-d12	15.35	264	260666	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.39	112	155561	16.87	ng	-0.01
7) Phenol-d6	6.43	99	196021	16.45	ng	-0.01
23) Nitrobenzene-d5	7.36	82	109989	10.85	ng	-0.02
41) 2,4,6-Tribromophenol	10.63	330	35333	16.65	ng	-0.01
44) 2-Fluorobiphenyl	9.16	172	208268	14.40	ng	-0.01
78) Terphenyl-d14	12.90	244	158550	11.40	ng	-0.01
Target Compounds						Qvalue
70) Phenanthrene	11.35	178	105308	4.53	ng	97
74) Fluoranthene	12.53	202	165684	7.04	ng	79
77) Pyrene	12.75	202	194833	7.42	ng	99
79) Butylbenzylphthalate	13.38	149	25412	2.17	ng	# 84
80) Benzo(a)anthracene	13.94	228	92642	4.65	ng	94
82) Chrysene	13.98	228	75604m	3.83	ng	
83) Bis(2-ethylhexyl)phthalate	13.94	149	40164	3.00	ng	97
85) Indeno(1,2,3-cd)pyrene	16.69	276	33693	2.22	ng	# 100
87) Benzo(b)fluoranthene	14.95	252	84667m	5.18	ng	
88) Benzo(k)fluoranthene	14.96	252	39409m	2.77	ng	
89) Benzo(a)pyrene	15.29	252	63957	4.62	ng	96
91) Benzo(g,h,i)perylene	17.10	276	35164	2.94	ng	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF070116\
Data File : BF088579.D
Acq On : 1 Jul 2016 17:17
Operator : UM/SJ
Sample : H3808-21 5X
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SB-11-COMP

Manual Integrations
APPROVED

sohil
7/5/2016 7:31:33 PM

Quant Time: Jul 01 18:03:46 2016
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF063016.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jun 30 18:01:55 2016
Response via : Initial Calibration

