


Data Path : Z:\HPCHEM1\BNA_F\DATA\BF070116\
 Data File : BF088587.D
 Acq On : 1 Jul 2016 21:41
 Operator : UM/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF063016.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Jun 30 18:01:55 2016



AutoFind: Scans 874, 875, 876; Background Corrected with Scan 871

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	43.5	14379	PASS
68	69	0.00	2	1.4	206	PASS
69	198	0.00	100	44.9	14858	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	55.1	18220	PASS
197	198	0.00	2	0.7	229	PASS
198	198	100	100	100.0	33083	PASS
199	198	5	9	6.6	2185	PASS
275	198	10	60	24.4	8061	PASS
365	198	1	100	3.0	981	PASS
441	442	0.01	24	14.9	4175	PASS
442	198	50	100	84.8	28050	PASS
443	442	15	24	18.7	5245	PASS