

Data Path : Z:\HPCHEM1\BNA_F\Data\BF070117\
 Data File : BF096377.D
 Acq On : 1 Jul 2017 14:24
 Operator : SJ/MA
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF070117

Quant Time: Jul 03 02:00:08 2017
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF070117.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jul 03 01:43:12 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	98	0.00
2	1,4-Dioxane	0.643	0.581	9.6	85	0.00
3	Pyridine	1.763	1.618	8.2	91	0.00
4	n-Nitrosodimethylamine	0.870	0.804	7.6	89	0.00
5 S	2-Fluorophenol	1.355	1.152	15.0	84	0.00
6	Aniline	2.153	2.135	0.8	98	0.00
7 S	Phenol-d6	1.666	1.635	1.9	97	0.00
8	2-Chlorophenol	1.445	1.413	2.2	97	0.00
9	Benzaldehyde	1.043	1.048	-0.5	98	0.00
10 C	Phenol	1.898	1.866	1.7	97	0.00
11	bis(2-Chloroethyl)ether	1.467	1.460	0.5	97	0.00
12	1,3-Dichlorobenzene	1.515	1.491	1.6	98	0.00
13 C	1,4-Dichlorobenzene	1.514	1.488	1.7	95	0.00
14	1,2-Dichlorobenzene	1.390	1.359	2.2	98	0.00
15	Benzyl Alcohol	1.114	1.114	0.0	98	0.00
16	2,2'-oxybis(1-Chloropropane	2.982	3.015	-1.1	98	0.00
17	2-Methylphenol	1.152	1.149	0.3	97	0.00
18	Hexachloroethane	0.549	0.544	0.9	96	0.00
19 P	n-Nitroso-di-n-propylamine	0.992	0.971	2.1	97	0.00
20	3+4-Methylphenols	1.284	1.271	1.0	96	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	99	0.00
22	Acetophenone	0.437	0.414	5.3	98	0.00
23 S	Nitrobenzene-d5	0.333	0.332	0.3	98	0.00
24	Nitrobenzene	0.349	0.351	-0.6	99	0.00
25	Isophorone	0.645	0.640	0.8	98	0.00
26 C	2-Nitrophenol	0.185	0.189	-2.2	98	0.00
27	2,4-Dimethylphenol	0.315	0.303	3.8	96	0.00
28	bis(2-Chloroethoxy)methane	0.426	0.423	0.7	98	0.00
29 C	2,4-Dichlorophenol	0.271	0.268	1.1	98	0.00
30	1,2,4-Trichlorobenzene	0.256	0.250	2.3	97	0.00
31	Naphthalene	0.944	0.914	3.2	97	0.00
32	Benzoic acid	0.264	0.240	9.1	87	0.00
33	4-Chloroaniline	0.392	0.384	2.0	97	0.00
34 C	Hexachlorobutadiene	0.131	0.125	4.6	95	0.00
35	Caprolactam	0.094	0.090	4.3	99	0.00
36 C	4-Chloro-3-methylphenol	0.300	0.289	3.7	98	0.00
37	2-Methylnaphthalene	0.601	0.572	4.8	97	0.00
38 I	Acenaphthene-d10	1.000	1.000	0.0	90	0.00
39	1,2,4,5-Tetrachlorobenzene	0.519	0.545	-5.0	95	0.00
40 P	Hexachlorocyclopentadiene	0.253	0.286	-13.0	96	0.00
41 S	2,4,6-Tribromophenol	0.156	0.169	-8.3	98	0.00
42 C	2,4,6-Trichlorophenol	0.411	0.429	-4.4	93	0.00
43	2,4,5-Trichlorophenol	0.406	0.436	-7.4	96	0.00
44 S	2-Fluorobiphenyl	1.170	1.250	-6.8	95	0.00

Data Path : Z:\HPCHEM1\BNA_F\Data\BF070117\
 Data File : BF096377.D
 Acq On : 1 Jul 2017 14:24
 Operator : SJ/MA
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF070117

Quant Time: Jul 03 02:00:08 2017
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF070117.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jul 03 01:43:12 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	1.714	1.786	-4.2	95	0.00
46	2-Chloronaphthalene	1.277	1.346	-5.4	95	0.00
47	2-Nitroaniline	0.465	0.497	-6.9	94	0.00
48	Acenaphthylene	2.024	1.984	2.0	89	0.00
49	Dimethylphthalate	1.493	1.467	1.7	90	0.00
50	2,6-Dinitrotoluene	0.330	0.336	-1.8	89	0.00
51 C	Acenaphthene	1.247	1.252	-0.4	91	0.00
52	3-Nitroaniline	0.412	0.402	2.4	88	0.00
53 P	2,4-Dinitrophenol	0.175	0.185	-5.7	90	0.00
54	Dibenzofuran	1.647	1.657	-0.6	90	0.00
55 P	4-Nitrophenol	0.341	0.342	-0.3	87	0.00
56	2,4-Dinitrotoluene	0.418	0.444	-6.2	92	0.00
57	Fluorene	1.163	1.284	-10.4	98	0.00
58	2,3,4,6-Tetrachlorophenol	0.281	0.309	-10.0	98	0.00
59	Diethylphthalate	1.411	1.513	-7.2	97	0.00
60	4-Chlorophenyl-phenylether	0.506	0.540	-6.7	97	0.00
61	4-Nitroaniline	0.374	0.413	-10.4	98	0.00
62	Azobenzene	1.365	1.506	-10.3	98	0.00
63 I	Phenanthrene-d10	1.000	1.000	0.0	99	0.00
64	4,6-Dinitro-2-methylphenol	0.138	0.145	-5.1	98	0.00
65 c	n-Nitrosodiphenylamine	0.702	0.690	1.7	98	0.00
66	4-Bromophenyl-phenylether	0.195	0.195	0.0	99	0.00
67	Hexachlorobenzene	0.213	0.212	0.5	99	0.00
68	Atrazine	0.200	0.194	3.0	97	0.00
69 C	Pentachlorophenol	0.126	0.136	-7.9	100	0.00
70	Phenanthrene	1.081	1.032	4.5	97	0.00
71	Anthracene	1.102	1.063	3.5	97	0.00
72	Carbazole	1.108	1.077	2.8	98	0.00
73	Di-n-butylphthalate	1.342	1.312	2.2	98	0.00
74 C	Fluoranthene	1.060	1.013	4.4	96	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	95	0.00
76	Benzidine	0.960	0.937	2.4	92	0.00
77	Pyrene	1.605	1.605	0.0	96	0.00
78 S	Terphenyl-d14	0.936	0.935	0.1	99	0.00
79	Butylbenzylphthalate	0.813	0.844	-3.8	98	0.00
80	Benzo(a)anthracene	1.158	1.166	-0.7	98	0.00
81	3,3'-Dichlorobenzidine	0.476	0.477	-0.2	94	0.00
82	Chrysene	1.213	1.226	-1.1	97	0.00
83	Bis(2-ethylhexyl)phthalate	0.907	0.909	-0.2	97	0.00
84 c	Di-n-octyl phthalate	1.786	1.821	-2.0	97	0.00
85	Indeno(1,2,3-cd)pyrene	0.957	0.933	2.5	98	0.00
86 I	Perylene-d12	1.000	1.000	0.0	95	0.00
87	Benzo(b)fluoranthene	1.253	1.210	3.4	90	0.00

Data Path : Z:\HPCHEM1\BNA_F\Data\BF070117\
Data File : BF096377.D
Acq On : 1 Jul 2017 14:24
Operator : SJ/MA
Sample : SSTDICV040
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
ICVBF070117

Quant Time: Jul 03 02:00:08 2017
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF070117.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jul 03 01:43:12 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	1.121	1.190	-6.2	105	0.00
89 C	Benzo(a)pyrene	1.119	1.126	-0.6	96	0.00
90	Dibenzo(a,h)anthracene	0.880	0.882	-0.2	98	0.00
91	Benzo(g,h,i)perylene	0.912	0.904	0.9	97	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0