

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF070124\
 Data File : BF138395.D
 Acq On : 01 Jul 2024 10:57
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jul 01 12:20:38 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF062824.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jun 29 02:44:30 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	111	0.00
2	1,4-Dioxane	40.000	42.376	-5.9	117	0.00
3	Pyridine	40.000	44.046	-10.1	119	0.00
4	n-Nitrosodimethylamine	40.000	41.944	-4.9	114	0.00
5 S	2-Fluorophenol	80.000	81.004	-1.3	113	0.00
6	Aniline	40.000	42.565	-6.4	120	0.00
7 S	Phenol-d6	80.000	81.452	-1.8	115	0.00
8	2-Chlorophenol	40.000	41.157	-2.9	116	0.00
9	Benzaldehyde	40.000	40.381	-1.0	117	0.00
10 C	Phenol	40.000	41.319	-3.3	117	0.00
11	bis(2-Chloroethyl)ether	40.000	42.144	-5.4	118	0.00
12	1,3-Dichlorobenzene	40.000	40.025	-0.1	112	0.00
13 C	1,4-Dichlorobenzene	40.000	39.988	0.0	113	0.00
14	1,2-Dichlorobenzene	40.000	40.035	-0.1	113	0.00
15	Benzyl Alcohol	40.000	41.273	-3.2	114	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	41.361	-3.4	116	0.00
17	2-Methylphenol	40.000	41.151	-2.9	116	0.00
18	Hexachloroethane	40.000	41.102	-2.8	113	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	39.401	1.5	112	0.00
20	3+4-Methylphenols	40.000	41.517	-3.8	115	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	114	0.00
22	Acetophenone	40.000	38.365	4.1	113	0.00
23 S	Nitrobenzene-d5	80.000	79.817	0.2	114	0.00
24	Nitrobenzene	40.000	40.317	-0.8	115	0.00
25	Isophorone	40.000	39.596	1.0	115	0.00
26 C	2-Nitrophenol	40.000	41.536	-3.8	113	0.00
27	2,4-Dimethylphenol	40.000	40.487	-1.2	116	0.00
28	bis(2-Chloroethoxy)methane	40.000	40.248	-0.6	116	0.00
29 C	2,4-Dichlorophenol	40.000	39.422	1.4	112	0.00
30	1,2,4-Trichlorobenzene	40.000	38.426	3.9	111	0.00
31	Naphthalene	40.000	38.868	2.8	113	0.00
32	Benzoic acid	40.000	42.333	-5.8	124	0.00
33	4-Chloroaniline	40.000	40.285	-0.7	117	0.00
34 C	Hexachlorobutadiene	40.000	38.455	3.9	111	0.00
35	Caprolactam	40.000	41.211	-3.0	118	0.00
36 C	4-Chloro-3-methylphenol	40.000	40.292	-0.7	117	0.00
37	2-Methylnaphthalene	40.000	38.538	3.7	113	0.00
38	1-Methylnaphthalene	40.000	38.646	3.4	113	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	114	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	37.777	5.6	110	0.00
41 P	Hexachlorocyclopentadiene	40.000	38.046	4.9	102	0.00
42 S	2,4,6-Tribromophenol	80.000	78.793	1.5	112	0.00
43 C	2,4,6-Trichlorophenol	40.000	39.268	1.8	111	0.00
44	2,4,5-Trichlorophenol	40.000	39.076	2.3	111	0.00
45 S	2-Fluorobiphenyl	80.000	73.536	8.1	110	0.00
46	1,1'-Biphenyl	40.000	38.223	4.4	111	0.00
47	2-Chloronaphthalene	40.000	38.505	3.7	112	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF070124\
 Data File : BF138395.D
 Acq On : 01 Jul 2024 10:57
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jul 01 12:20:38 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF062824.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jun 29 02:44:30 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	41.510	-3.8	118	0.00
49	Acenaphthylene	40.000	38.783	3.0	114	0.00
50	Dimethylphthalate	40.000	38.878	2.8	115	0.00
51	2,6-Dinitrotoluene	40.000	39.977	0.1	114	0.00
52 C	Acenaphthene	40.000	39.013	2.5	113	0.00
53	3-Nitroaniline	40.000	40.657	-1.6	117	0.00
54 P	2,4-Dinitrophenol	40.000	38.343	4.1	117	0.00
55	Dibenzofuran	40.000	38.779	3.1	114	0.00
56 P	4-Nitrophenol	40.000	42.331	-5.8	116	0.00
57	2,4-Dinitrotoluene	40.000	41.150	-2.9	114	0.00
58	Fluorene	40.000	36.279	9.3	112	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	40.476	-1.2	113	0.00
60	Diethylphthalate	40.000	39.039	2.4	115	0.00
61	4-Chlorophenyl-phenylether	40.000	36.784	8.0	113	0.00
62	4-Nitroaniline	40.000	41.824	-4.6	117	0.00
63	Azobenzene	40.000	40.307	-0.8	117	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	118	0.00
65	4,6-Dinitro-2-methylphenol	40.000	38.281	4.3	116	0.00
66 c	n-Nitrosodiphenylamine	40.000	37.769	5.6	115	0.00
67	4-Bromophenyl-phenylether	40.000	37.988	5.0	115	0.00
68	Hexachlorobenzene	40.000	36.623	8.4	112	0.00
69	Atrazine	40.000	40.478	-1.2	114	0.00
70 C	Pentachlorophenol	40.000	41.393	-3.5	114	0.00
71	Phenanthrene	40.000	37.916	5.2	114	0.00
72	Anthracene	40.000	38.467	3.8	116	0.00
73	Carbazole	40.000	39.524	1.2	117	0.00
74	Di-n-butylphthalate	40.000	39.843	0.4	118	0.00
75 C	Fluoranthene	40.000	40.243	-0.6	118	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	119	0.00
77	Benzydine	40.000	11.916	70.2#	176	0.00
78	Pyrene	40.000	40.180	-0.4	119	0.00
79 S	Terphenyl-d14	80.000	79.045	1.2	115	0.00
80	Butylbenzylphthalate	40.000	43.504	-8.8	123	0.00
81	Benzo(a)anthracene	40.000	39.613	1.0	119	0.00
82	3,3'-Dichlorobenzidine	40.000	37.376	6.6	119	0.00
83	Chrysene	40.000	39.683	0.8	121	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	41.869	-4.7	122	0.00
85 c	Di-n-octyl phthalate	40.000	38.604	3.5	116	0.00
86 I	Perylene-d12	20.000	20.000	0.0	112	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	37.911	5.2	103	0.00
88	Benzo(b)fluoranthene	40.000	39.298	1.8	116	0.00
89	Benzo(k)fluoranthene	40.000	39.860	0.4	112	0.00
90 C	Benzo(a)pyrene	40.000	39.316	1.7	113	0.00
91	Dibenzo(a,h)anthracene	40.000	38.067	4.8	101	0.00
92	Benzo(g,h,i)perylene	40.000	38.923	2.7	104	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF070124\
Data File : BF138395.D
Acq On : 01 Jul 2024 10:57
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Jul 01 12:20:38 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF062824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Jun 29 02:44:30 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0