

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF070125\
 Data File : BF142961.D
 Acq On : 01 Jul 2025 21:04
 Operator : RC/JU
 Sample : Q2458-02 5X
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 TP-55

Quant Time: Jul 02 01:04:37 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF061925.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 24 05:28:21 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.875	152	38271	20.000	ng	0.00
21) Naphthalene-d8	8.157	136	133925	20.000	ng	-0.01
39) Acenaphthene-d10	9.916	164	74002	20.000	ng	0.00
64) Phenanthrene-d10	11.404	188	149997	20.000	ng	0.00
76) Chrysene-d12	14.051	240	111978	20.000	ng	0.00
86) Perylene-d12	15.545	264	84459	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.493	112	36283	15.784	ng	0.00
7) Phenol-d6	6.504	99	38447	14.177	ng	0.00
23) Nitrobenzene-d5	7.434	82	21999	9.010	ng	-0.01
42) 2,4,6-Tribromophenol	10.704	330	13298	17.462	ng	-0.01
45) 2-Fluorobiphenyl	9.233	172	56127	9.964	ng	-0.01
79) Terphenyl-d14	12.986	244	78959	9.743	ng	-0.01

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF070125\
Data File : BF142961.D
Acq On : 01 Jul 2025 21:04
Operator : RC/JU
Sample : Q2458-02 5X
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
TP-55

Quant Time: Jul 02 01:04:37 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF061925.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jun 24 05:28:21 2025
Response via : Initial Calibration

