

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF070215\
 Data File : BF080336.D
 Acq On : 3 Jul 2015 10:55
 Operator : TP/IZ
 Sample : G2873-04
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 BOTTOMMNW-2-070115

Quant Time: Jul 03 23:27:56 2015
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF070115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 03 22:57:58 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.24	152	31459	20.00	ng	0.00
21) Naphthalene-d8	8.53	136	114749	20.00	ng	0.00
38) Acenaphthene-d10	10.29	164	54754	20.00	ng	-0.01
63) Phenanthrene-d10	11.80	188	117980	20.00	ng	0.00
75) Chrysene-d12	14.74	240	132306	20.00	ng	0.02
86) Perylene-d12	16.58	264	126519	20.00	ng	0.03
System Monitoring Compounds						
5) 2-Fluorophenol	5.85	112	291237	160.32	ng	0.00
7) Phenol-d6	6.84	99	350620	143.78	ng	0.00
23) Nitrobenzene-d5	7.80	82	201029	96.46	ng	0.00
41) 2,4,6-Tribromophenol	11.09	330	95874	183.89	ng	0.00
44) 2-Fluorobiphenyl	9.61	172	446056	117.25	ng	0.00
78) Terphenyl-d14	13.48	244	531592	94.11	ng	0.01
Target Compounds						
49) Dimethylphthalate	10.00	163	77208	18.46	ng	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF070215\
Data File : BF080336.D
Acq On : 3 Jul 2015 10:55
Operator : TP/IZ
Sample : G2873-04
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
BOTTOMNW-2-070115

Quant Time: Jul 03 23:27:56 2015
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF070115.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jul 03 22:57:58 2015
Response via : Initial Calibration

