


Data Path : Z:\HPCHEM1\BNA_F\DATA\BF070216\
 Data File : BF088610.D
 Acq On : 2 Jul 2016 11:49
 Operator : UM/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF063016.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Jun 30 18:01:55 2016



AutoFind: Scans 873, 874, 875; Background Corrected with Scan 870

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	42.0	16975	PASS
68	69	0.00	2	1.5	274	PASS
69	198	0.00	100	44.0	17816	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	55.4	22399	PASS
197	198	0.00	2	0.7	273	PASS
198	198	100	100	100.0	40462	PASS
199	198	5	9	6.3	2557	PASS
275	198	10	60	25.1	10143	PASS
365	198	1	100	3.0	1231	PASS
441	442	0.01	24	15.4	5210	PASS
442	198	50	100	83.8	33925	PASS
443	442	15	24	19.2	6513	PASS