

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF070225\
 Data File : BF142972.D
 Acq On : 02 Jul 2025 15:47
 Operator : RC/JU
 Sample : Q2452-08
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 EP-2

Quant Time: Jul 02 16:12:55 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF061925.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 24 05:28:21 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.875	152	53635	20.000	ng	0.00
21) Naphthalene-d8	8.157	136	201811	20.000	ng	-0.01
39) Acenaphthene-d10	9.910	164	111172	20.000	ng	-0.01
64) Phenanthrene-d10	11.404	188	192487	20.000	ng	0.00
76) Chrysene-d12	14.045	240	111458	20.000	ng	-0.01
86) Perylene-d12	15.539	264	108899	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.498	112	373890	116.062	ng	0.00
7) Phenol-d6	6.504	99	419762	110.443	ng	0.00
23) Nitrobenzene-d5	7.434	82	289698	78.738	ng	-0.01
42) 2,4,6-Tribromophenol	10.704	330	151763	132.656	ng	-0.01
45) 2-Fluorobiphenyl	9.233	172	640480	75.683	ng	-0.01
79) Terphenyl-d14	12.992	244	661126	81.957	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF070225\
Data File : BF142972.D
Acq On : 02 Jul 2025 15:47
Operator : RC/JU
Sample : Q2452-08
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
EP-2

Quant Time: Jul 02 16:12:55 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF061925.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jun 24 05:28:21 2025
Response via : Initial Calibration

