


Data Path : Z:\HPCHEM1\BNA_F\DATA\BF070316\
 Data File : BF088675.D
 Acq On : 4 Jul 2016 4:40
 Operator : UM/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF063016.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Jul 05 17:15:08 2016



AutoFind: Scans 872, 873, 874; Background Corrected with Scan 869

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	40.5	17062	PASS
68	69	0.00	2	1.4	256	PASS
69	198	0.00	100	42.5	17911	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	54.6	22987	PASS
197	198	0.00	2	0.5	229	PASS
198	198	100	100	100.0	42121	PASS
199	198	5	9	6.6	2776	PASS
275	198	10	60	24.5	10321	PASS
365	198	1	100	2.9	1218	PASS
441	442	0.01	24	15.2	5156	PASS
442	198	50	100	80.5	33922	PASS
443	442	15	24	19.2	6528	PASS