


Data Path : Z:\HPCHEM1\BNA_F\DATA\BF070516\
 Data File : BF088697.D
 Acq On : 5 Jul 2016 11:51
 Operator : UM/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF063016.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Jul 06 16:39:34 2016



AutoFind: Scans 871, 872, 873; Background Corrected with Scan 868

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	39.3	21195	PASS
68	69	0.00	2	1.5	343	PASS
69	198	0.00	100	42.3	22832	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	54.3	29264	PASS
197	198	0.00	2	0.3	187	PASS
198	198	100	100	100.0	53926	PASS
199	198	5	9	6.6	3556	PASS
275	198	10	60	26.5	14307	PASS
365	198	1	100	3.5	1877	PASS
441	442	0.01	24	15.8	8244	PASS
442	198	50	100	97.0	52294	PASS
443	442	15	24	18.5	9656	PASS