


Data Path : Z:\HPCHEM1\BNA_F\DATA\BF070516\
 Data File : BF088705.D
 Acq On : 5 Jul 2016 16:14
 Operator : UM/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF063016.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Jul 06 16:39:34 2016



AutoFind: Scans 871, 872, 873; Background Corrected with Scan 868

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	41.9	21736	PASS
68	69	0.00	2	1.5	346	PASS
69	198	0.00	100	44.8	23256	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	55.7	28921	PASS
197	198	0.00	2	0.2	109	PASS
198	198	100	100	100.0	51927	PASS
199	198	5	9	6.8	3553	PASS
275	198	10	60	26.0	13498	PASS
365	198	1	100	3.6	1869	PASS
441	442	0.01	24	15.8	8135	PASS
442	198	50	100	98.9	51370	PASS
443	442	15	24	18.8	9666	PASS