

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF070518\
 Data File : BF107128.D
 Acq On : 5 Jul 2018 15:51
 Operator : JU/SJ
 Sample : J3853-07
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 B-6

Manual Integrations
 APPROVED

Sohil
 7/6/2018 8:45:16 AM

Quant Time: Jul 05 16:34:41 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF061918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 29 14:38:29 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.95	152	57026	20.00	ng	-0.01
21) Naphthalene-d8	8.23	136	211651	20.00	ng	-0.02
38) Acenaphthene-d10	10.00	164	98191	20.00	ng	-0.01
63) Phenanthrene-d10	11.49	188	161382	20.00	ng	-0.01
75) Chrysene-d12	14.15	240	151983	20.00	ng	-0.02
86) Perylene-d12	15.69	264	136586	20.00	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.60	112	348393	103.45	ng	0.00
7) Phenol-d6	6.60	99	432263	102.78	ng	-0.02
23) Nitrobenzene-d5	7.51	82	267352	70.20	ng	-0.02
41) 2,4,6-Tribromophenol	10.79	330	113058	99.34	ng	-0.01
44) 2-Fluorobiphenyl	9.31	172	467388	79.64	ng	-0.02
78) Terphenyl-d14	13.07	244	388942	61.99	ng	-0.01
Target Compounds						Qvalue
31) Naphthalene	8.25	128	22618	2.286	ng	99
37) 2-Methylnaphthalene	8.95	142	22840	3.482	ng	97
49) Dimethylphthalate	9.70	163	96822	13.147	ng	# 98
70) Phenanthrene	11.51	178	42456	5.454	ng	98
74) Fluoranthene	12.71	202	32230	3.757	ng	97
77) Pyrene	12.94	202	26362	2.630	ng	98
82) Chrysene	14.17	228	22278	2.614	ng	# 95
87) Benzo(b)fluoranthene	15.23	252	23787m	2.804	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF070518\
Data File : BF107128.D
Acq On : 5 Jul 2018 15:51
Operator : JU/SJ
Sample : J3853-07
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
B-6

Quant Time: Jul 05 16:34:41 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF061918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 29 14:38:29 2018
Response via : Initial Calibration

Manual Integrations
APPROVED

Sohil
7/6/2018 8:45:16 AM

