

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF070522\
 Data File : BF129131.D
 Acq On : 05 Jul 2022 19:03
 Operator : CG\JU
 Sample : N3564-27
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 WC-16-2-4

Quant Time: Jul 06 02:25:59 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF062922.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 29 17:05:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.957	152	327179	20.000	ng	0.00
21) Naphthalene-d8	8.239	136	1223315	20.000	ng	0.00
39) Acenaphthene-d10	9.998	164	586379	20.000	ng	0.00
64) Phenanthrene-d10	11.492	188	920879	20.000	ng	0.00
76) Chrysene-d12	14.139	240	788482	20.000	ng	#-0.01
86) Perylene-d12	15.668	264	793852	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.592	112	2157555	111.487	ng	0.00
7) Phenol-d6	6.581	99	2383994	99.188	ng	0.00
23) Nitrobenzene-d5	7.522	82	1816969	88.592	ng	0.00
42) 2,4,6-Tribromophenol	10.792	330	770517	120.846	ng	0.00
45) 2-Fluorobiphenyl	9.316	172	3257382	88.869	ng	0.00
79) Terphenyl-d14	13.080	244	3141991	82.751	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF070522\
Data File : BF129131.D
Acq On : 05 Jul 2022 19:03
Operator : CG\JU
Sample : N3564-27
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
WC-16-2-4

Quant Time: Jul 06 02:25:59 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF062922.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 29 17:05:31 2022
Response via : Initial Calibration

