

Data Path : Z:\svoasrv\HPCHEM1\BNA\_F\Data\BF070524\  
 Data File : BF138472.D  
 Acq On : 05 Jul 2024 17:04  
 Operator : CG\JU  
 Sample : P3135-07RE  
 Misc :  
 ALS Vial : 12 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 OK-628-COMP-02RE

Quant Time: Jul 05 17:24:53 2024  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_F\Methods\8270-BF062824.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Sat Jun 29 02:44:30 2024  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon | Response | Conc   | Units | Dev(Min) |
|-----------------------------|--------|------|----------|--------|-------|----------|
| -----                       |        |      |          |        |       |          |
| Internal Standards          |        |      |          |        |       |          |
| 1) 1,4-Dichlorobenzene-d4   | 6.869  | 152  | 56626    | 20.000 | ng    | 0.00     |
| 21) Naphthalene-d8          | 8.151  | 136  | 221593   | 20.000 | ng    | 0.00     |
| 39) Acenaphthene-d10        | 9.904  | 164  | 115221   | 20.000 | ng    | -0.01    |
| 64) Phenanthrene-d10        | 11.392 | 188  | 176391   | 20.000 | ng    | 0.00     |
| 76) Chrysene-d12            | 14.027 | 240  | 116449   | 20.000 | ng    | 0.00     |
| 86) Perylene-d12            | 15.498 | 264  | 83443    | 20.000 | ng    | 0.00     |
| System Monitoring Compounds |        |      |          |        |       |          |
| 5) 2-Fluorophenol           | 5.487  | 112  | 339395   | 99.371 | ng    | -0.01    |
| 7) Phenol-d6                | 6.504  | 99   | 444866   | 96.691 | ng    | -0.01    |
| 23) Nitrobenzene-d5         | 7.428  | 82   | 277318   | 65.011 | ng    | -0.02    |
| 42) 2,4,6-Tribromophenol    | 10.692 | 330  | 83561    | 88.765 | ng    | -0.01    |
| 45) 2-Fluorobiphenyl        | 9.228  | 172  | 489254   | 67.577 | ng    | 0.00     |
| 79) Terphenyl-d14           | 12.974 | 244  | 413179   | 48.625 | ng    | 0.00     |

|                  |        |
|------------------|--------|
| Target Compounds | Qvalue |
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA\_F\Data\BF070524\  
Data File : BF138472.D  
Acq On : 05 Jul 2024 17:04  
Operator : CG\JU  
Sample : P3135-07RE  
Misc :  
ALS Vial : 12 Sample Multiplier: 1

Instrument :  
BNA\_F  
ClientSampleId :  
OK-628-COMP-02RE

Quant Time: Jul 05 17:24:53 2024  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_F\Methods\8270-BF062824.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Sat Jun 29 02:44:30 2024  
Response via : Initial Calibration

