





Method Name: Z:\HPCHEM1\BNA_F\METHODS\8270-BF070615.M
Calibration Table Last Updated: Tue Jul 07 03:05:40 2015


Hexachlorocyclopentadiene

Response Ratio




Resp Ratio = $2.89 \times 10^{-1} \times \text{Amt} - 1.02 \times 10^{-1}$
Coef of Det (r^2) = 0.992 Curve Fit: Linear

Method Name: Z:\HPCHEM1\BNA_F\METHODS\8270-BF070615.M
Calibration Table Last Updated: Tue Jul 07 03:05:40 2015



Method Name: Z:\HPCHEM1\BNA_F\METHODS\8270-BF070615.M
Calibration Table Last Updated: Tue Jul 07 03:05:40 2015

4,6-Dinitro-2-methylphenol



Resp Ratio = $1.31 \times 10^{-1} \times \text{Amt} - 3.46 \times 10^{-2}$
Coef of Det (r^2) = 0.993 Curve Fit: Linear

Method Name: Z:\HPCHEM1\BNA_F\METHODS\8270-BF070615.M
Calibration Table Last Updated: Tue Jul 07 03:05:40 2015