

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF070615\
 Data File : BF080360.D
 Acq On : 6 Jul 2015 22:51
 Operator : TP/IZ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jul 07 05:39:19 2015
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF070615.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 07 03:05:40 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	98	0.00
2	1,4-Dioxane	0.516	0.539	-4.5	101	-0.01
3	Pyridine	1.312	1.414	-7.8	111	0.00
4	n-Nitrosodimethylamine	0.617	0.692	-12.2	110	-0.01
5 S	2-Fluorophenol	1.092	1.089	0.3	98	0.00
6	Aniline	2.004	2.035	-1.5	99	0.00
7 S	Phenol-d6	1.448	1.414	2.3	99	0.00
8	2-Chlorophenol	1.266	1.270	-0.3	98	0.00
9	Benzaldehyde	0.805	0.827	-2.7	101	0.00
10 C	Phenol	1.570	1.501	4.4	98	0.00
11	bis(2-Chloroethyl)ether	1.184	1.152	2.7	100	0.00
12	1,3-Dichlorobenzene	1.554	1.536	1.2	96	0.00
13 C	1,4-Dichlorobenzene	1.451	1.484	-2.3	101	-0.01
14	1,2-Dichlorobenzene	1.469	1.475	-0.4	99	0.00
15	Benzyl Alcohol	1.120	1.171	-4.6	99	0.00
16	2,2'-oxybis(1-Chloropropane	1.915	1.917	-0.1	100	0.00
17	2-Methylphenol	0.981	0.981	0.0	98	0.00
18	Hexachloroethane	0.466	0.464	0.4	96	0.00
19 P	n-Nitroso-di-n-propylamine	0.939	0.945	-0.6	99	0.00
20	3+4-Methylphenols	1.368	1.415	-3.4	99	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	101	0.00
22	Acetophenone	0.461	0.464	-0.7	97	0.00
23 S	Nitrobenzene-d5	0.343	0.348	-1.5	97	0.00
24	Nitrobenzene	0.344	0.339	1.5	100	0.00
25	Isophorone	0.660	0.670	-1.5	101	0.00
26 C	2-Nitrophenol	0.149	0.152	-2.0	97	0.00
27	2,4-Dimethylphenol	0.293	0.290	1.0	100	0.00
28	bis(2-Chloroethoxy)methane	0.410	0.406	1.0	98	0.00
29 C	2,4-Dichlorophenol	0.260	0.270	-3.8	103	0.00
30	1,2,4-Trichlorobenzene	0.309	0.299	3.2	97	0.00
31	Naphthalene	1.026	1.023	0.3	99	0.00
32	Benzoic acid	0.143	0.171	-19.6	109	0.00
33	4-Chloroaniline	0.409	0.413	-1.0	103	0.00
34 C	Hexachlorobutadiene	0.188	0.184	2.1	97	0.00
35	Caprolactam	0.080	0.081	-1.3	100	0.00
36 C	4-Chloro-3-methylphenol	0.269	0.261	3.0	99	0.00
37	2-Methylnaphthalene	0.654	0.642	1.8	97	0.00
38 I	Acenaphthene-d10	1.000	1.000	0.0	103	0.00
39	1,2,4,5-Tetrachlorobenzene	0.634	0.616	2.8	98	0.00
40 P	Hexachlorocyclopentadiene	0.223	0.220	1.3	97	0.00
41 S	2,4,6-Tribromophenol	0.209	0.212	-1.4	99	0.00
42 C	2,4,6-Trichlorophenol	0.418	0.415	0.7	96	0.00
43	2,4,5-Trichlorophenol	0.487	0.488	-0.2	100	0.00
44 S	2-Fluorobiphenyl	1.482	1.467	1.0	100	0.00

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF070615\
 Data File : BF080360.D
 Acq On : 6 Jul 2015 22:51
 Operator : TP/IZ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jul 07 05:39:19 2015
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF070615.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 07 03:05:40 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	1.775	1.748	1.5	100	0.00
46	2-Chloronaphthalene	1.354	1.348	0.4	99	0.00
47	2-Nitroaniline	0.379	0.396	-4.5	101	0.00
48	Acenaphthylene	2.253	2.193	2.7	99	0.00
49	Dimethylphthalate	1.606	1.456	9.3	99	0.00
50	2,6-Dinitrotoluene	0.336	0.343	-2.1	100	0.00
51 C	Acenaphthene	1.349	1.367	-1.3	102	0.00
52	3-Nitroaniline	0.378	0.384	-1.6	97	0.00
53 P	2,4-Dinitrophenol	0.120	0.118	1.7	92	0.00
54	Dibenzofuran	1.918	1.922	-0.2	103	0.00
55 P	4-Nitrophenol	0.277	0.299	-7.9	99	0.00
56	2,4-Dinitrotoluene	0.432	0.434	-0.5	100	0.00
57	Fluorene	1.651	1.650	0.1	99	0.00
58	2,3,4,6-Tetrachlorophenol	0.390	0.395	-1.3	99	0.00
59	Diethylphthalate	1.604	1.579	1.6	98	0.00
60	4-Chlorophenyl-phenylether	0.731	0.727	0.5	99	0.00
61	4-Nitroaniline	0.367	0.397	-8.2	100	0.00
62	Azobenzene	1.523	1.513	0.7	98	0.00
63 I	Phenanthrene-d10	1.000	1.000	0.0	100	0.00
64	4,6-Dinitro-2-methylphenol	0.106	0.105	0.9	94	0.00
65 c	n-Nitrosodiphenylamine	0.637	0.647	-1.6	100	0.00
66	4-Bromophenyl-phenylether	0.229	0.227	0.9	98	0.00
67	Hexachlorobenzene	0.240	0.246	-2.5	101	0.00
68	Atrazine	0.190	0.200	-5.3	105	0.00
69 C	Pentachlorophenol	0.119	0.100	16.0	88	0.00
70	Phenanthrene	1.200	1.186	1.2	104	0.00
71	Anthracene	1.176	1.171	0.4	97	0.00
72	Carbazole	1.097	1.136	-3.6	105	0.00
73	Di-n-butylphthalate	1.255	1.238	1.4	98	0.00
74 C	Fluoranthene	1.289	1.247	3.3	98	-0.01
75 I	Chrysene-d12	1.000	1.000	0.0	100	-0.01
76	Benzidine	0.507	0.550	-8.5	102	0.00
77	Pyrene	1.210	1.274	-5.3	103	0.00
78 S	Terphenyl-d14	0.857	0.914	-6.7	106	0.00
79	Butylbenzylphthalate	0.483	0.530	-9.7	100	-0.01
80	Benzo(a)anthracene	1.205	1.221	-1.3	99	-0.01
81	3,3'-Dichlorobenzidine	0.400	0.417	-4.2	97	-0.01
82	Chrysene	1.128	1.159	-2.7	98	-0.01
83	Bis(2-ethylhexyl)phthalate	0.736	0.794	-7.9	97	-0.01
84 c	Di-n-octyl phthalate	1.215	1.299	-6.9	97	-0.01
85	Indeno(1,2,3-cd)pyrene	1.326	1.354	-2.1	97	0.00
86 I	Perylene-d12	1.000	1.000	0.0	97	0.00
87	Benzo(b)fluoranthene	1.257	1.213	3.5	87	0.00

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF070615\
Data File : BF080360.D
Acq On : 6 Jul 2015 22:51
Operator : TP/IZ
Sample : SSTDCCC040
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Jul 07 05:39:19 2015
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF070615.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 07 03:05:40 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	1.190	1.301	-9.3	112	0.00
89 C	Benzo(a)pyrene	1.149	1.187	-3.3	98	-0.01
90	Dibenzo(a,h)anthracene	1.132	1.157	-2.2	97	0.00
91	Benzo(g,h,i)perylene	1.148	1.178	-2.6	97	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0