


Data Path : Z:\HPCHEM1\BNA F\DATA\BF070716\
 Data File : BF088805.D
 Acq On : 8 Jul 2016 4:29
 Operator : UM/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF063016.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Jul 05 17:15:08 2016



AutoFind: Scans 869, 870, 871; Background Corrected with Scan 866

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	36.8	6279	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	40.0	6830	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	53.3	9102	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	17075	PASS
199	198	5	9	6.1	1041	PASS
275	198	10	60	27.0	4602	PASS
365	198	1	100	2.7	453	PASS
441	442	0.01	24	15.7	2300	PASS
442	198	50	100	85.5	14607	PASS
443	442	15	24	19.3	2818	PASS