

Data Path : Z:\HPCHEM1\BNA_F\Data\BF070717\
 Data File : BF096550.D
 Acq On : 7 Jul 2017 11:06
 Operator : SJ/MA
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jul 07 13:00:48 2017
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF070117.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jul 05 13:27:10 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	74	0.00
2	1,4-Dioxane	0.643	0.655	-1.9	72	0.00
3	Pyridine	1.763	1.548	12.2	65	0.00
4	n-Nitrosodimethylamine	0.870	0.840	3.4	70	0.00
5 S	2-Fluorophenol	1.355	1.263	6.8	70	0.00
6	Aniline	2.153	1.970	8.5	68	0.00
7 S	Phenol-d6	1.666	1.563	6.2	70	0.00
8	2-Chlorophenol	1.445	1.414	2.1	73	0.00
9	Benzaldehyde	1.043	0.895	14.2	63	0.00
10 C	Phenol	1.898	1.728	9.0	68	0.00
11	bis(2-Chloroethyl)ether	1.467	1.363	7.1	68	0.00
12	1,3-Dichlorobenzene	1.515	1.501	0.9	75	0.00
13 C	1,4-Dichlorobenzene	1.514	1.538	-1.6	74	0.00
14	1,2-Dichlorobenzene	1.390	1.410	-1.4	77	0.00
15	Benzyl Alcohol	1.114	1.076	3.4	71	0.00
16	2,2'-oxybis(1-Chloropropane	2.982	2.838	4.8	70	0.00
17	2-Methylphenol	1.152	1.132	1.7	72	0.00
18	Hexachloroethane	0.549	0.554	-0.9	74	0.00
19 P	n-Nitroso-di-n-propylamine	0.992	0.940	5.2	71	0.00
20	3+4-Methylphenols	1.284	1.311	-2.1	75	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	83	0.00
22	Acetophenone	0.437	0.386	11.7	75	0.00
23 S	Nitrobenzene-d5	0.333	0.292	12.3	72	0.00
24	Nitrobenzene	0.349	0.303	13.2	71	0.00
25	Isophorone	0.645	0.547	15.2	70	0.00
26 C	2-Nitrophenol	0.185	0.172	7.0	74	0.00
27	2,4-Dimethylphenol	0.315	0.281	10.8	74	0.00
28	bis(2-Chloroethoxy)methane	0.426	0.371	12.9	71	0.00
29 C	2,4-Dichlorophenol	0.271	0.269	0.7	82	0.00
30	1,2,4-Trichlorobenzene	0.256	0.258	-0.8	83	0.00
31	Naphthalene	0.944	0.931	1.4	82	0.00
32	Benzoic acid	0.264	0.228	13.6	69	0.00
33	4-Chloroaniline	0.392	0.383	2.3	81	0.00
34 C	Hexachlorobutadiene	0.131	0.133	-1.5	84	0.00
35	Caprolactam	0.094	0.086	8.5	79	0.00
36 C	4-Chloro-3-methylphenol	0.300	0.270	10.0	76	0.00
37	2-Methylnaphthalene	0.601	0.586	2.5	82	0.00
38 I	Acenaphthene-d10	1.000	1.000	0.0	85	0.00
39	1,2,4,5-Tetrachlorobenzene	0.519	0.528	-1.7	87	0.00
40 P	Hexachlorocyclopentadiene	0.253	0.239	5.5	75	0.00
41 S	2,4,6-Tribromophenol	0.156	0.165	-5.8	90	0.00
42 C	2,4,6-Trichlorophenol	0.411	0.379	7.8	78	0.00
43	2,4,5-Trichlorophenol	0.406	0.403	0.7	83	0.00
44 S	2-Fluorobiphenyl	1.170	1.196	-2.2	85	0.00

Data Path : Z:\HPCHEM1\BNA_F\Data\BF070717\
 Data File : BF096550.D
 Acq On : 7 Jul 2017 11:06
 Operator : SJ/MA
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jul 07 13:00:48 2017
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF070117.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jul 05 13:27:10 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	1.714	1.681	1.9	84	0.00
46	2-Chloronaphthalene	1.277	1.229	3.8	82	0.00
47	2-Nitroaniline	0.465	0.401	13.8	72	0.00
48	Acenaphthylene	2.024	1.840	9.1	78	0.00
49	Dimethylphthalate	1.493	1.355	9.2	78	0.00
50	2,6-Dinitrotoluene	0.330	0.305	7.6	77	0.00
51 C	Acenaphthene	1.247	1.140	8.6	78	0.00
52	3-Nitroaniline	0.412	0.399	3.2	82	0.00
53 P	2,4-Dinitrophenol	0.175	0.168	4.0	77	0.00
54	Dibenzofuran	1.647	1.516	8.0	78	0.00
55 P	4-Nitrophenol	0.341	0.318	6.7	76	0.00
56	2,4-Dinitrotoluene	0.418	0.396	5.3	77	0.00
57	Fluorene	1.163	1.181	-1.5	85	0.00
58	2,3,4,6-Tetrachlorophenol	0.281	0.280	0.4	84	0.00
59	Diethylphthalate	1.411	1.279	9.4	78	0.00
60	4-Chlorophenyl-phenylether	0.506	0.497	1.8	84	0.00
61	4-Nitroaniline	0.374	0.364	2.7	82	0.00
62	Azobenzene	1.365	1.309	4.1	80	0.00
63 I	Phenanthrene-d10	1.000	1.000	0.0	85	0.00
64	4,6-Dinitro-2-methylphenol	0.138	0.138	0.0	80	0.00
65 c	n-Nitrosodiphenylamine	0.702	0.671	4.4	82	0.00
66	4-Bromophenyl-phenylether	0.195	0.186	4.6	81	0.00
67	Hexachlorobenzene	0.213	0.202	5.2	81	0.00
68	Atrazine	0.200	0.188	6.0	81	0.00
69 C	Pentachlorophenol	0.126	0.118	6.3	75	0.00
70	Phenanthrene	1.081	1.045	3.3	85	0.00
71	Anthracene	1.102	1.056	4.2	83	0.00
72	Carbazole	1.108	1.012	8.7	79	0.00
73	Di-n-butylphthalate	1.342	1.190	11.3	76	0.00
74 C	Fluoranthene	1.060	0.979	7.6	79	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	91	0.00
76	Benzidine	0.960	0.816	15.0	77	0.00
77	Pyrene	1.605	1.403	12.6	80	0.00
78 S	Terphenyl-d14	0.936	0.831	11.2	85	0.00
79	Butylbenzylphthalate	0.813	0.746	8.2	83	0.00
80	Benzo(a)anthracene	1.158	1.117	3.5	90	0.00
81	3,3'-Dichlorobenzidine	0.476	0.492	-3.4	93	0.00
82	Chrysene	1.213	1.197	1.3	91	0.00
83	Bis(2-ethylhexyl)phthalate	0.907	0.875	3.5	90	0.00
84 c	Di-n-octyl phthalate	1.786	1.765	1.2	91	0.00
85	Indeno(1,2,3-cd)pyrene	0.957	1.008	-5.3	102	0.00
86 I	Perylene-d12	1.000	1.000	0.0	105	0.00
87	Benzo(b)fluoranthene	1.253	1.130	9.8	92	0.00

Data Path : Z:\HPCHEM1\BNA_F\Data\BF070717\
Data File : BF096550.D
Acq On : 7 Jul 2017 11:06
Operator : SJ/MA
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Jul 07 13:00:48 2017
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF070117.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jul 05 13:27:10 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	1.121	1.138	-1.5	110	0.00
89 C	Benzo(a)pyrene	1.119	1.070	4.4	101	0.00
90	Dibenzo(a,h)anthracene	0.880	0.832	5.5	102	0.00
91	Benzo(g,h,i)perylene	0.912	0.865	5.2	102	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0