

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF070721\
 Data File : BF124541.D
 Acq On : 7 Jul 2021 21:50
 Operator : JU/CG
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICSVBF070721

Quant Time: Jul 08 07:56:54 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF070721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jul 08 07:49:42 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	114	0.00
2	1,4-Dioxane	40.000	43.259	-8.1	119	0.02
3	Pyridine	40.000	41.499	-3.7	116	0.02
4	n-Nitrosodimethylamine	40.000	40.030	-0.1	108	0.02
5 S	2-Fluorophenol	80.000	80.363	-0.5	114	0.00
6	Aniline	40.000	39.087	2.3	112	0.00
7 S	Phenol-d6	80.000	80.666	-0.8	112	0.00
8	2-Chlorophenol	40.000	40.703	-1.8	113	0.00
9	Benzaldehyde	40.000	37.041	7.4	115	0.00
10 C	Phenol	40.000	40.391	-1.0	113	0.00
11	bis(2-Chloroethyl)ether	40.000	41.340	-3.4	117	0.00
12	1,3-Dichlorobenzene	40.000	39.348	1.6	112	0.00
13 C	1,4-Dichlorobenzene	40.000	40.342	-0.9	113	0.00
14	1,2-Dichlorobenzene	40.000	40.400	-1.0	115	0.00
15	Benzyl Alcohol	40.000	40.695	-1.7	111	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	40.040	-0.1	112	0.00
17	2-Methylphenol	40.000	39.802	0.5	111	0.00
18	Hexachloroethane	40.000	41.483	-3.7	112	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	40.062	-0.2	111	0.00
20	3+4-Methylphenols	40.000	40.426	-1.1	118	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	115	0.00
22	Acetophenone	40.000	39.309	1.7	112	0.00
23 S	Nitrobenzene-d5	80.000	81.353	-1.7	114	0.00
24	Nitrobenzene	40.000	39.562	1.1	109	0.00
25	Isophorone	40.000	39.176	2.1	112	0.00
26 C	2-Nitrophenol	40.000	42.489	-6.2	118	0.00
27	2,4-Dimethylphenol	40.000	40.581	-1.5	115	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.071	2.3	113	0.00
29 C	2,4-Dichlorophenol	40.000	40.366	-0.9	111	0.00
30	1,2,4-Trichlorobenzene	40.000	38.571	3.6	111	0.00
31	Naphthalene	40.000	39.871	0.3	114	0.00
32	Benzoic acid	40.000	40.381	-1.0	118	0.01
33	4-Chloroaniline	40.000	39.692	0.8	110	0.00
34 C	Hexachlorobutadiene	40.000	39.589	1.0	111	0.00
35	Caprolactam	40.000	37.622	5.9	119	0.00
36 C	4-Chloro-3-methylphenol	40.000	41.617	-4.0	115	0.00
37	2-Methylnaphthalene	40.000	39.497	1.3	112	0.00
38	1-Methylnaphthalene	40.000	39.460	1.3	114	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	114	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	40.888	-2.2	114	0.00
41 P	Hexachlorocyclopentadiene	40.000	39.723	0.7	110	0.00
42 S	2,4,6-Tribromophenol	80.000	84.006	-5.0	118	0.00
43 C	2,4,6-Trichlorophenol	40.000	41.830	-4.6	115	0.00
44	2,4,5-Trichlorophenol	40.000	41.553	-3.9	121	0.00
45 S	2-Fluorobiphenyl	80.000	79.759	0.3	113	0.00
46	1,1'-Biphenyl	40.000	39.503	1.2	113	0.00
47	2-Chloronaphthalene	40.000	39.784	0.5	112	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF070721\
 Data File : BF124541.D
 Acq On : 7 Jul 2021 21:50
 Operator : JU/CG
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF070721

Quant Time: Jul 08 07:56:54 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF070721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jul 08 07:49:42 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	45.621	-14.1	119	0.00
49	Acenaphthylene	40.000	39.506	1.2	112	0.00
50	Dimethylphthalate	40.000	40.920	-2.3	116	0.00
51	2,6-Dinitrotoluene	40.000	44.425	-11.1	112	0.00
52 C	Acenaphthene	40.000	41.055	-2.6	117	0.00
53	3-Nitroaniline	40.000	41.810	-4.5	116	0.00
54 P	2,4-Dinitrophenol	40.000	43.725	-9.3	125	0.00
55	Dibenzofuran	40.000	39.308	1.7	112	0.00
56 P	4-Nitrophenol	40.000	43.828	-9.6	118	0.00
57	2,4-Dinitrotoluene	40.000	44.018	-10.0	119	0.00
58	Fluorene	40.000	38.344	4.1	112	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	41.631	-4.1	117	0.00
60	Diethylphthalate	40.000	40.068	-0.2	114	0.00
61	4-Chlorophenyl-phenylether	40.000	41.401	-3.5	117	0.00
62	4-Nitroaniline	40.000	40.516	-1.3	115	0.00
63	Azobenzene	40.000	40.958	-2.4	119	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	115	0.00
65	4,6-Dinitro-2-methylphenol	40.000	42.397	-6.0	116	0.00
66 c	n-Nitrosodiphenylamine	40.000	40.859	-2.1	114	0.00
67	4-Bromophenyl-phenylether	40.000	40.970	-2.4	115	0.00
68	Hexachlorobenzene	40.000	40.282	-0.7	109	0.00
69	Atrazine	40.000	40.459	-1.1	113	0.00
70 C	Pentachlorophenol	40.000	43.487	-8.7	123	0.00
71	Phenanthrene	40.000	40.441	-1.1	113	0.00
72	Anthracene	40.000	40.342	-0.9	113	0.00
73	Carbazole	40.000	40.215	-0.5	113	0.00
74	Di-n-butylphthalate	40.000	40.705	-1.8	113	0.00
75 C	Fluoranthene	40.000	40.119	-0.3	110	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	111	0.00
77	Benzydine	40.000	36.826	7.9	103	0.00
78	Pyrene	40.000	39.439	1.4	110	0.00
79 S	Terphenyl-d14	80.000	79.564	0.5	114	0.00
80	Butylbenzylphthalate	40.000	39.831	0.4	110	0.00
81	Benzo(a)anthracene	40.000	39.416	1.5	110	0.00
82	3,3'-Dichlorobenzidine	40.000	40.827	-2.1	109	0.00
83	Chrysene	40.000	40.052	-0.1	110	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	40.542	-1.4	112	0.00
85 c	Di-n-octyl phthalate	40.000	39.691	0.8	109	0.00
86 I	Perylene-d12	20.000	20.000	0.0	111	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	43.091	-7.7	113	0.00
88	Benzo(b)fluoranthene	40.000	41.676	-4.2	113	0.00
89	Benzo(k)fluoranthene	40.000	38.484	3.8	97	0.00
90 C	Benzo(a)pyrene	40.000	40.165	-0.4	109	0.00
91	Dibenzo(a,h)anthracene	40.000	42.681	-6.7	114	0.00
92	Benzo(g,h,i)perylene	40.000	42.896	-7.2	113	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF070721\
Data File : BF124541.D
Acq On : 7 Jul 2021 21:50
Operator : JU/CG
Sample : SSTDICV040
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
ICVBF070721

Quant Time: Jul 08 07:56:54 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF070721.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 08 07:49:42 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0