


Data Path : Z:\HPCHEM1\BNA_F\DATA\BF070815\
 Data File : BF080384.D
 Acq On : 8 Jul 2015 12:41
 Operator : TP/UM
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF070615.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Jul 07 03:05:40 2015



AutoFind: Scans 871, 872, 873; Background Corrected with Scan 868

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	37.1	5749	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	34.9	5419	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	44.7	6933	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	15508	PASS
199	198	5	9	6.5	1015	PASS
275	198	10	60	25.3	3919	PASS
365	198	1	100	2.3	360	PASS
441	442	0.01	24	15.4	1798	PASS
442	198	50	100	75.1	11644	PASS
443	442	15	24	18.7	2182	PASS