


Data Path : V:\HPCHEM1\BNA F\DATA\BF070816\
 Data File : BF088829.D
 Acq On : 8 Jul 2016 17:25
 Operator : UM/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF063016.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Jul 08 18:51:20 2016



AutoFind: Scans 867, 868, 869; Background Corrected with Scan 864

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	40.0	9491	PASS
68	69	0.00	2	1.3	131	PASS
69	198	0.00	100	43.5	10314	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	55.8	13243	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	23736	PASS
199	198	5	9	6.8	1611	PASS
275	198	10	60	26.6	6315	PASS
365	198	1	100	3.6	845	PASS
441	442	0.01	24	15.6	3625	PASS
442	198	50	100	97.7	23188	PASS
443	442	15	24	18.9	4374	PASS