

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF070819\
 Data File : BF115428.D
 Acq On : 9 Jul 2019 1:37
 Operator : HP/JU
 Sample : K3681-01
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PILEA-07-01-19

Quant Time: Jul 09 06:01:47 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF070519.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 05 14:57:35 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.89	152	190517	20.00	ng	0.00
21) Naphthalene-d8	8.17	136	755024	20.00	ng	0.00
39) Acenaphthene-d10	9.93	164	375430	20.00	ng	0.00
64) Phenanthrene-d10	11.41	188	670574	20.00	ng	0.00
76) Chrysene-d12	14.04	240	417516	20.00	ng	-0.03
87) Perylene-d12	15.51	264	452668	20.00	ng	-0.05
System Monitoring Compounds						
5) 2-Fluorophenol	5.52	112	1084035	87.91	ng	0.00
7) Phenol-d6	6.52	99	1428296	91.08	ng	0.00
23) Nitrobenzene-d5	7.45	82	892960	69.88	ng	0.00
42) 2,4,6-Tribromophenol	10.72	330	338948	81.65	ng	0.00
45) 2-Fluorobiphenyl	9.25	172	1501069	70.16	ng	0.00
79) Terphenyl-d14	12.99	244	1390665	68.20	ng	-0.01
Target Compounds						
32) Benzoic acid	7.85	122	5135	6.587	ng	88
50) Dimethylphthalate	9.64	163	228125	8.096	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF070819\
Data File : BF115428.D
Acq On : 9 Jul 2019 1:37
Operator : HP/JU
Sample : K3681-01
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PILEA-07-01-19

Quant Time: Jul 09 06:01:47 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF070519.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jul 05 14:57:35 2019
Response via : Initial Calibration

