


Data Path : Z:\SVOASRV\HPCHEM1\BNA_F\DATA\BF070819\
 Data File : BF115410.D
 Acq On : 8 Jul 2019 17:24
 Operator : HP/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF070519.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Jul 05 14:57:35 2019



AutoFind: Scans 1629, 1630, 1631; Background Corrected with Scan 1624

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	32.1	52240	PASS
68	69	0.00	2	1.8	982	PASS
69	198	0.00	100	34.0	55365	PASS
70	69	0.00	2	0.8	418	PASS
127	198	10	80	47.1	76677	PASS
197	198	0.00	2	0.5	839	PASS
198	198	100	100	100.0	162754	PASS
199	198	5	9	6.7	10893	PASS
275	198	10	60	27.0	43965	PASS
365	198	1	100	3.4	5482	PASS
441	198	0.01	100	13.8	22497	PASS
442	198	50	100	91.1	148224	PASS
443	442	15	24	18.3	27114	PASS