




Data Path : Z:\HPCHEM1\BNA_F\Data\BF070916\
Data File : BF088863.D
Acq On : 9 Jul 2016 11:11
Operator : UM/SJ
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jul 11 07:48:08 2016
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF063016.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jul 08 18:51:20 2016
Response via : Initial Calibration



TIC: BF088863.D

(76) Benzidine




12.559min (-0.011) 0.00ng

response 280965

Ion	Exp%	Act%
184.00	100	100
185.00	15.60	14.43
183.00	11.60	12.01
0.00	0.00	0.00

Data Path : Z:\HPCHEM1\BNA_F\Data\BF070916\
Data File : BF088863.D
Acq On : 9 Jul 2016 11:11
Operator : UM/SJ
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jul 11 07:48:08 2016
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF063016.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jul 08 18:51:20 2016
Response via : Initial Calibration



TIC: BF088863.D

(69) Pentachlorophenol (C)

11.039min (-0.011) 0.00ng

response 45741

Ion	Exp%	Act%
265.70	100	100
268.00	66.00	64.35
264.00	62.00	62.34
0.00	0.00	0.00