

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF070919\
 Data File : BF115434.D
 Acq On : 9 Jul 2019 15:25
 Operator : HP/JU
 Sample : K3705-01
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 GW-1-NE-BOILER-PIT

Quant Time: Jul 09 16:38:19 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF070519.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 09 16:29:09 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.89	152	159320	20.00	ng	0.00
21) Naphthalene-d8	8.17	136	615810	20.00	ng	0.00
39) Acenaphthene-d10	9.93	164	314945	20.00	ng	0.00
64) Phenanthrene-d10	11.42	188	623528	20.00	ng	0.00
76) Chrysene-d12	14.18	240	478877	20.00	ng	-0.04
87) Perylene-d12	15.83	264	407252	20.00	ng	-0.05
System Monitoring Compounds						
5) 2-Fluorophenol	5.51	112	543208	52.68	ng	0.00
7) Phenol-d6	6.51	99	484148	36.92	ng	-0.02
23) Nitrobenzene-d5	7.45	82	998688	95.82	ng	0.00
42) 2,4,6-Tribromophenol	10.72	330	472076	135.57	ng	0.00
45) 2-Fluorobiphenyl	9.25	172	1716673	95.65	ng	0.00
79) Terphenyl-d14	13.03	244	2123066	90.77	ng	0.00
Target Compounds						
32) Benzoic acid	7.89	122	18920	8.502	ng	98
50) Dimethylphthalate	9.64	163	52960	2.240	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF070919\
Data File : BF115434.D
Acq On : 9 Jul 2019 15:25
Operator : HP/JU
Sample : K3705-01
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
GW-1-NE-BOILER-PIT

Quant Time: Jul 09 16:38:19 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF070519.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 09 16:29:09 2019
Response via : Initial Calibration

